DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Labora	atory Name :	Alliance Technical Group LLC	Client :	G Environmenta	al				
Project Location : NJ		Project Number :	Banker						
Laboratory Sample ID(s): Q3275 Sampling Date(s): 10/02/2025									
List DI	KQP Methods U	sed (e.g., 8260,8270, et Cetra) 300.0,6	6010D,SOP						
1	specified QA/Q explain any crit	tical method referenced in this laboratory C performance criteria followed, includin eria falling outside of acceptable guidelin Known Quality performance standards?	g the requirement to nes, as specified in th		7	Yes		No	
1A	Were the methor	od specified handling, preservation, and	holding time requiren	nents met?	$\overline{\mathbf{A}}$	Yes		No	
1B		Vas the EPH method conducted without respective DKQ methods)	significant modification	ons (see		Yes		No	✓ N/A
2		es received by the laboratory in a conditine associated chain-of-custody documen		at	V	Yes		No	
3	Were samples	received at an appropriate temperature (4±2° C)?		V	Yes		No	□ N/A
4	Were all QA/Q0 standards ach	C performance criteria specified in the Noieved?	JDEP DKQP			Yes	V	No	
5		ng limits specified or referenced on the cl to the laboratory prior to sample receipt?			V	Yes		No	
	b)Were these re	eporting limits met?			V	Yes		No	□ N/A
6	results reporte	tical method referenced in this laboratory ed for all constituents identified in the met e DKQP documents and/or site-specific (thod-specific analyte		V	Yes		No	
7	Are project-spe	ecific matrix spikes and/or laboratory dup	licates included in thi	s data set?		Yes	$\overline{\mathbf{A}}$	No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Cover Page

Order ID:	Q3275
-----------	-------

Project ID: Banker

Client: G Environmental

Lab Sample Number Client Sample Number Q3275-01 MW3 Q3275-02 MW4 Q3275-03 MW5 Q3275-04 MW1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :		
orginatare r	 Date:	10/13/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

CASE NARRATIVE

G Environmental Project Name: Banker Project # N/A

Order ID # Q3275

Test Name: Metals Group4, Anions Group1

A. Number of Samples and Date of Receipt:

4 Water samples were received on 10/02/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Metals Group4, Anions Group1. This data package contains results for Metals Group4(6010D), Anions Group1(300.0).

C. Analytical Techniques:

Metals Group4: The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

Wetchem: The analysis of Anions Group1 was based on method 300.0.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

 $Metals\ Group 4: Sample\ MW5\ was\ diluted\ due\ to\ high\ concentrations\ for\ Sodium.$ Wetchem: Sample\ MW3\ was\ diluted\ due\ to\ high\ concentrations\ for\ Chloride\ \&\ Sample

MW4 was diluted due to high concentrations for Chloride & Sample MW5 was diluted due to high concentrations for Chloride & Sample MW5DL was diluted due to high concentrations for Chloride & Sample MW1 was diluted due to high concentrations for Chloride.

The MS recoveries met the requirements for all compounds except following Wetchem: The Matrix Spike (MW1MS) analysis met criteria for all compounds except for Chloride due to sample matrix interference.

The MSD recoveries met the requirements for all compounds except following Wetchem: The Matrix Spike Duplicate (MW1MSD) analysis met criteria for all compounds except for Chloride due to sample matrix interference.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Duplicate analysis met criteria for all compound.

The Serial Dilution met the acceptable requirements.



284 Sheffield Street, Mountainside, NJ 7092, Phone: 908 789 8900, Fax: 908 789 8922

E. Additional Comments:

Metals Group4: In analytical sequence LB137466, The Result was outside of acceptance limit for Sodium of CCB05 and CCB06 but, no any sample associated under these CCBs.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.



DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following "Results Qualifiers" are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).		
U	Indicates the analyte was analyzed for, but not detected.		
ND	Indicates the analyte was analyzed for, but not detected		
E	Indicates the reported value is estimated because of the presence of interference		
M	Indicates Duplicate injection precision not met.		
N	Indicates the spiked sample recovery is not within control limits.		
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).		
*	Indicates that the duplicate analysis is not within control limits.		
+	Indicates the correlation coefficient for the MSA is less than 0.995.		
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.		
M	Method qualifiers "P" for ICP instrument "PM" for ICP when Microwave Digestion is used "CV" for Manual Cold Vapor AA "AV" for automated Cold Vapor AA "CA" for MIDI-Distillation Spectrophotometric "AS" for Semi –Automated Spectrophotometric "C" for Manual Spectrophotometric "T" for Titrimetric "NR" for analyte not required to be analyzed Indicates the analyte's concentration exceeds the calibrated range of the instrument for that specific analysis.		
Q	Indicates the LCS did not meet the control limits requirements		
Н	Sample Analysis Out Of Hold Time		





APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3275

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication, airbill, sample management lab chronicle, login page)	<u> </u>
Check chain-of-custody for proper relinquish/return of samples	<u>✓</u>
Is the chain of custody signed and complete	<u> </u>
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	<u> </u>
Collect information for each project id from server. Were all requirements followed	<u> </u>
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	<u>✓</u>
Do lab numbers and client Ids on cover page agree with the Chain of Custody	<u> </u>
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	<u> </u>
Do requested analyses on Chain of Custody agree with the log-in page	<u> </u>
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	<u>✓</u>
Were the samples received within hold time	<u>✓</u>
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	_
Does the case narrative summarize all QC failure?	_
All runlogs and manual integration are reviewed for requirements	
All manual calculations and /or hand notations verified	'

QA Review Signature: SOHIL JODHANI	Date:	10/13/2025
------------------------------------	-------	------------