

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137410

Review By	rubina	Review On	10/7/2025 8:23:35 AM
Supervise By	Iwona	Supervise On	10/7/2025 9:31:32 AM
SubDirectory	LB137410	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114677,WP114678,WP114679,WP114680,WP114681,WP114682,WP114683		
ICV Standard	WP114684		
CCV Standard	WP115067		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115068		
Chk Standard	WP114685,WP114686		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	STD1	STD1	CAL1	09/09/25 10:21	All standards, samples, and	RM/IZ	OK
2	STD2	STD2	CAL2	09/09/25 10:42	QC are filtered through	RM/IZ	OK
3	STD3	STD3	CAL3	09/09/25 11:25	0.45um, filter lot W3160	RM/IZ	OK
4	STD4	STD4	CAL4	09/09/25 11:47		RM/IZ	OK
5	STD5	STD5	CAL5	09/09/25 13:36		RM/IZ	OK
6	STD6	STD6	CAL6	09/09/25 13:58		RM/IZ	OK
7	STD7	STD7	CAL7	09/09/25 14:41		RM/IZ	OK
8	ICV1	ICV1	ICV	09/09/25 15:02		RM/IZ	OK
9	ICB1	ICB1	ICB	09/09/25 15:45		RM/IZ	OK
10	CCV1	CCV1	CCV	10/03/25 09:11		RM/IZ	OK
11	CCB1	CCB1	CCB	10/03/25 09:33		RM/IZ	OK
12	LB137410BLW	LB137410BLW	MB	10/03/25 09:54		RM/IZ	OK
13	LB137410BSW	LB137410BSW	LCS	10/03/25 10:16		RM/IZ	OK
14	Q3275-01	MW3	SAM	10/03/25 10:37	Cl is high, need dilution	RM/IZ	Dilution
15	Q3275-02	MW4	SAM	10/03/25 10:59	Cl is high, need dilution	RM/IZ	Dilution
16	Q3275-03	MW5	SAM	10/03/25 11:21	Cl is high, need dilution	RM/IZ	Dilution
17	Q3275-04	MW1	SAM	10/03/25 11:42	Cl is high, need dilution	RM/IZ	Dilution
18	Q3275-04MS	MW1MS	MS	10/03/25 12:04	9.5ml of sample, 0.5mL W3096	RM/IZ	OK

Instrument ID: IC-1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137410

Review By	rubina	Review On	10/7/2025 8:23:35 AM
Supervise By	Iwona	Supervise On	10/7/2025 9:31:32 AM
SubDirectory	LB137410	Test	Anions
STD. NAME	STD REF.#		
ICAL Standard	WP114677,WP114678,WP114679,WP114680,WP114681,WP114682,WP114683		
ICV Standard	WP114684		
CCV Standard	WP115067		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP115068		
Chk Standard	WP114685,WP114686		

19	Q3275-04MSD	MW1MSD	MSD	10/03/25 12:25	9.5ml of sample, 0.5mL W3096	RM/IZ	OK
20	Q3278-01	MW10	SAM	10/03/25 12:47		RM/IZ	OK
21	CCV2	CCV2	CCV	10/03/25 13:08		RM/IZ	OK
22	CCB2	CCB2	CCB	10/03/25 13:30		RM/IZ	OK
23	Q3275-01DL	MW3DL	SAM	10/03/25 13:51	100X for Cl	RM/IZ	Confirms
24	Q3275-02DL	MW4DL	SAM	10/03/25 14:13	100X for Cl	RM/IZ	OK
25	Q3275-03DL	MW5DL	SAM	10/03/25 14:36	500X For cl still High	RM/IZ	Dilution
26	Q3275-04DL	MW1DL	SAM	10/03/25 15:00	200X For Cl	RM/IZ	Confirms
27	Q3275-03DL2	MW5DL2	SAM	10/03/25 15:21	5000X For Cl	RM/IZ	Confirms
28	CCV3	CCV3	CCV	10/03/25 15:43		RM/IZ	OK
29	CCB3	CCB3	CCB	10/03/25 16:04		RM/IZ	OK