



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3275 **Order ID:** Q3275
Client: G Environmental **Project ID:** Banker

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : MW3 Q3275-01	MW3	Water	Sodium	200000		434	1000	ug/L
Client ID : MW4 Q3275-02	MW4	Water	Sodium	148000		434	1000	ug/L
Client ID : MW5 Q3275-03	MW5	Water	Sodium	11200000	D	4340	10000	ug/L
Client ID : MW1 Q3275-04	MW1	Water	Sodium	345000		434	1000	ug/L



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	10/02/25
Project:	Banker	Date Received:	10/02/25
Client Sample ID:	MW3	SDG No.:	Q3275
Lab Sample ID:	Q3275-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	200000	1	434		1000	ug/L	10/03/25 10:20	10/08/25 15:47	6010D	SW3010

Color Before:	light Brown	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	10/02/25
Project:	Banker	Date Received:	10/02/25
Client Sample ID:	MW4	SDG No.:	Q3275
Lab Sample ID:	Q3275-02	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	148000	1	434		1000	ug/L	10/03/25 10:20	10/08/25 15:51	6010D	SW3010

Color Before:	Brown	Clarity Before:	Cloudy	Texture:
Color After:	light Brown	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	10/02/25
Project:	Banker	Date Received:	10/02/25
Client Sample ID:	MW5	SDG No.:	Q3275
Lab Sample ID:	Q3275-03	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	11200000	D	10	4340	10000	ug/L	10/03/25 10:20	10/09/25 17:14	6010D	SW3010

Color Before: light Brown

Clarity Before: Clear

Texture:

Color After: Colorless

Clarity After: Clear

Artifacts:

Comments: Metals Group4

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

Client:	G Environmental	Date Collected:	10/02/25
Project:	Banker	Date Received:	10/02/25
Client Sample ID:	MW1	SDG No.:	Q3275
Lab Sample ID:	Q3275-04	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-23-5	Sodium	345000	1	434		1000	ug/L	10/03/25 10:20	10/09/25 13:09	6010D	SW3010

Color Before:	light Brown	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group4			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	868	+/-1000	U	2000	P	10/08/2025	12:36	LB137466

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	868	+/-1000	U	2000	P	10/08/2025	13:16	LB137466
CCB02	Sodium	868	+/-1000	U	2000	P	10/08/2025	14:18	LB137466
CCB03	Sodium	868	+/-1000	U	2000	P	10/08/2025	15:13	LB137466
CCB04	Sodium	868	+/-1000	U	2000	P	10/08/2025	16:15	LB137466
CCB05	Sodium	1530	+/-1000	J*	2000	P	10/08/2025	17:06	LB137466
CCB06	Sodium	1100	+/-1000	J*	2000	P	10/08/2025	17:57	LB137466

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Sodium	868	+/-1000	U	2000	P	10/09/2025	11:59	LB137484

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Sodium	868	+/-1000	U	2000	P	10/09/2025	12:46	LB137484
CCB02	Sodium	868	+/-1000	U	2000	P	10/09/2025	14:20	LB137484
CCB03	Sodium	868	+/-1000	U	2000	P	10/09/2025	15:30	LB137484
CCB04	Sodium	868	+/-1000	U	2000	P	10/09/2025	16:41	LB137484
CCB05	Sodium	868	+/-1000	U	2000	P	10/09/2025	17:31	LB137484
CCB06	Sodium	868	+/-1000	U	2000	P	10/09/2025	18:20	LB137484
CCB07	Sodium	868	+/-1000	U	2000	P	10/09/2025	18:33	LB137484

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3275

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169978BL		WATER		Batch Number:	PB169978		Prep Date:	10/03/2025	
	Sodium	434	<500	U	1000	P	10/08/2025	15:30	LB137466



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19000	20000	95	90 - 110	P	10/08/2025	12:24	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1770	2000	89	80 - 120	P	10/08/2025	12:32	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

Contract: GENV01

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3275

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	25900	25000	104	90 - 110	P	10/08/2025	13:12	LB137466
CCV02	Sodium	24100	25000	96	90 - 110	P	10/08/2025	14:14	LB137466
CCV03	Sodium	22600	25000	90	90 - 110	P	10/08/2025	15:04	LB137466
CCV04	Sodium	26100	25000	104	90 - 110	P	10/08/2025	16:11	LB137466
CCV05	Sodium	26900	25000	108	90 - 110	P	10/08/2025	17:02	LB137466
CCV06	Sodium	26000	25000	104	90 - 110	P	10/08/2025	17:53	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19600	20000	98	90 - 110	P	10/09/2025	11:44	LB137484

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1610	2000	80	80 - 120	P	10/09/2025	11:49	LB137484

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

Contract: GENV01

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

SDG No.: Q3275

Lab Code: ACE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	24400	25000	98	90 - 110	P	10/09/2025	12:34	LB137484
CCV02	Sodium	23500	25000	94	90 - 110	P	10/09/2025	14:16	LB137484
CCV03	Sodium	23400	25000	94	90 - 110	P	10/09/2025	15:26	LB137484
CCV04	Sodium	24300	25000	97	90 - 110	P	10/09/2025	16:26	LB137484
CCV05	Sodium	23100	25000	92	90 - 110	P	10/09/2025	17:27	LB137484
CCV06	Sodium	26600	25000	106	90 - 110	P	10/09/2025	18:16	LB137484
CCV07	Sodium	26300	25000	105	90 - 110	P	10/09/2025	18:28	LB137484



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Sodium	1840	2000	92	65 - 135	P	10/08/2025	12:41	LB137466
CRI01	Sodium	1610	2000	80	65 - 135	P	10/09/2025	12:04	LB137484

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Sodium	60.4			0	0	10/08/2025	12:45	LB137466
ICSAB01	Sodium	40.2			0	0	10/08/2025	12:50	LB137466
ICSA01	Sodium	-178			0	0	10/09/2025	12:09	LB137484
ICSAB01	Sodium	-302			0	0	10/09/2025	12:19	LB137484



METAL QC DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3275</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	381000		371000		1500	705		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client:	G Environmental	level:	low	sdg no.:	Q3275
contract:	GENV01	lab code:	ACE		
matrix:	Water	sample id:	Q3268-01	client id:	WATER-TREATMENT-DISCHARGEMSD
Percent Solids for Sample:	NA	Spiked ID:	Q3268-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Sodium	ug/L	75 - 125	376000		371000		1500	347		P

Metals
- 5b -

Client:

G Environmental

Contract:

GENV01

Matrix:

Sample ID:

SDG No.:

Q3275

Lab Code:

ACE

Client ID:

Level:

LOW

Spiked ID:

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

- A
- B
- C
- D
- E
- F
- G
- H
- I
- J

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3275</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3268-01</u>	Client ID:	<u>WATER-TREATMENT-DISCHARGEDU</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3268-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	371000		370000		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>G Environmental</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3275</u>
Contract:	<u>GENV01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3268-01MS</u>	Client ID:	<u>WATER-TREATMENT-DISCHARGE</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3268-01MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Sodium	ug/L	20	381000		376000		1		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169978BS Sodium	ug/L	1500	1570		105	80 - 120	P

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

WATER-TREATMENT-DISCHARGEL

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

Lb No.: lb137466

Lab Sample ID : Q3268-01L

SDG No.: Q3275

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C	Serial Dilution Result (S) C	% Differ- ence	Q	M
Sodium	371000	361000	2		P

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q3275

Instrument id number: _____

Method: _____

Run number: LB137466

Start date: 10/08/2025

End date: 10/08/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1200	Na
S1	S1	1	1204	Na
S2	S2	1	1208	Na
S3	S3	1	1212	Na
S4	S4	1	1216	Na
S5	S5	1	1220	Na
ICV01	ICV01	1	1224	Na
LLICV01	LLICV01	1	1232	Na
ICB01	ICB01	1	1236	Na
CRI01	CRI01	1	1241	Na
ICSA01	ICSA01	1	1245	Na
ICSAB01	ICSAB01	1	1250	Na
CCV01	CCV01	1	1312	Na
CCB01	CCB01	1	1316	Na
CCV02	CCV02	1	1414	Na
CCB02	CCB02	1	1418	Na
Q3268-01DUP	WATER-TREATMENT-DISCH/	1	1447	Na
Q3268-01L	WATER-TREATMENT-DISCH/	5	1452	Na
Q3268-01MS	WATER-TREATMENT-DISCH/	1	1456	Na
Q3268-01MSD	WATER-TREATMENT-DISCH/	1	1500	Na
CCV03	CCV03	1	1504	Na
CCB03	CCB03	1	1513	Na
PB169978BL	PB169978BL	1	1530	Na
PB169978BS	PB169978BS	1	1534	Na
Q3275-01	MW3	1	1547	Na
Q3275-02	MW4	1	1551	Na
CCV04	CCV04	1	1611	Na
CCB04	CCB04	1	1615	Na
CCV05	CCV05	1	1702	Na
CCB05	CCB05	1	1706	Na
CCV06	CCV06	1	1753	Na
CCB06	CCB06	1	1757	Na

metals
- 14 -
ANALYSIS RUN LOG

Client: G Environmental

Contract: GENV01

Lab code: ACE

Sdg no.: Q3275

Instrument id number: _____

Method: _____

Run number: LB137484

Start date: 10/09/2025

End date: 10/09/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1110	Na
S1	S1	1	1115	Na
S2	S2	1	1119	Na
S3	S3	1	1123	Na
S4	S4	1	1127	Na
S5	S5	1	1131	Na
ICV01	ICV01	1	1144	Na
LLICV01	LLICV01	1	1149	Na
ICB01	ICB01	1	1159	Na
CRI01	CRI01	1	1204	Na
ICSA01	ICSA01	1	1209	Na
ICSAB01	ICSAB01	1	1219	Na
CCV01	CCV01	1	1234	Na
CCB01	CCB01	1	1246	Na
Q3275-04	MW1	1	1309	Na
CCV02	CCV02	1	1416	Na
CCB02	CCB02	1	1420	Na
CCV03	CCV03	1	1526	Na
CCB03	CCB03	1	1530	Na
CCV04	CCV04	1	1626	Na
CCB04	CCB04	1	1641	Na
Q3275-03	MW5	10	1714	Na
CCV05	CCV05	1	1727	Na
CCB05	CCB05	1	1731	Na
CCV06	CCV06	1	1816	Na
CCB06	CCB06	1	1820	Na
CCV07	CCV07	1	1828	Na
CCB07	CCB07	1	1833	Na



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		As	Ba	Be	Cd	Co
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Sodium	589.592	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000

LAB CHRONICLE

OrderID:	Q3275	OrderDate:	10/2/2025 3:39:00 PM
Client:	G Environmental	Project:	Banker
Contact:	Gary Landis	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3275-01	MW3	Water	Metals Group4	6010D	10/02/25	10/03/25	10/08/25	10/02/25
Q3275-02	MW4	Water	Metals Group4	6010D	10/02/25	10/03/25	10/08/25	10/02/25
Q3275-03	MW5	Water	Metals Group4	6010D	10/02/25	10/03/25	10/09/25	10/02/25
Q3275-04	MW1	Water	Metals Group4	6010D	10/02/25	10/03/25	10/09/25	10/02/25



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Method: _____

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB169978							
PB169978BL	PB169978BL	MB	WATER	10/03/2025	50.0	25.0	
PB169978BS	PB169978BS	LCS	WATER	10/03/2025	50.0	25.0	
Q3268-01DUP	WATER-TREATMENT-DISCHARGEDUP	DUP	WATER	10/03/2025	50.0	25.0	
Q3268-01MS	WATER-TREATMENT-DISCHARGEMS	MS	WATER	10/03/2025	50.0	25.0	
Q3268-01MSD	WATER-TREATMENT-DISCHARGEMSD	MSD	WATER	10/03/2025	50.0	25.0	
Q3275-01	MW3	SAM	WATER	10/03/2025	50.0	25.0	
Q3275-02	MW4	SAM	WATER	10/03/2025	50.0	25.0	
Q3275-03	MW5	SAM	WATER	10/03/2025	50.0	25.0	
Q3275-04	MW1	SAM	WATER	10/03/2025	50.0	25.0	

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/08/25 12:00		Jaswal	OK
2	S1	S1	CAL2	10/08/25 12:04		Jaswal	OK
3	S2	S2	CAL3	10/08/25 12:08		Jaswal	OK
4	S3	S3	CAL4	10/08/25 12:12		Jaswal	OK
5	S4	S4	CAL5	10/08/25 12:16		Jaswal	OK
6	S5	S5	CAL6	10/08/25 12:20		Jaswal	OK
7	ICV01	ICV01	ICV	10/08/25 12:24		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/08/25 12:32		Jaswal	OK
9	ICB01	ICB01	ICB	10/08/25 12:36		Jaswal	OK
10	CRI01	CRI01	CRDL	10/08/25 12:41		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/08/25 12:45		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/08/25 12:50		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/08/25 12:54		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/08/25 12:59		Jaswal	OK
15	CCV01	CCV01	CCV	10/08/25 13:12		Jaswal	OK
16	CCB01	CCB01	CCB	10/08/25 13:16		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/08/25 13:25		Jaswal	OK
18	LR-3	LR-3	HIGH STD	10/08/25 13:29		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	Q3295-03	EO-02-10062025	SAM	10/08/25 13:34		Jaswal	OK
20	Q3295-01	EO-01-10062025	SAM	10/08/25 13:38		Jaswal	OK
21	Q3295-01L	EO-01-10062025L	SD	10/08/25 13:46		Jaswal	OK
22	Q3295-01MS	EO-01-10062025MS	MS	10/08/25 13:50		Jaswal	OK
23	Q3295-01MSD	EO-01-10062025MSD	MSD	10/08/25 13:54		Jaswal	OK
24	Q3295-01A	EO-01-10062025A	PS	10/08/25 13:58	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
25	LR-1	LR-1	HIGH STD	10/08/25 14:02		Jaswal	OK
26	Q3295-01DUP	EO-01-10062025DUP	DUP	10/08/25 14:08		Jaswal	OK
27	CCV02	CCV02	CCV	10/08/25 14:14		Jaswal	OK
28	CCB02	CCB02	CCB	10/08/25 14:18		Jaswal	OK
29	Q3296-01	SU-04-10062025	SAM	10/08/25 14:23		Jaswal	OK
30	Q3293-01	CONC-STORAGE-PA	SAM	10/08/25 14:27		Jaswal	OK
31	Q3294-01	HD-02-10062025	SAM	10/08/25 14:31		Jaswal	OK
32	Q3297-01	TP1	SAM	10/08/25 14:35		Jaswal	OK
33	Q3297-11	TP2	SAM	10/08/25 14:39		Jaswal	OK
34	Q3268-01	WATER-TREATMENT	SAM	10/08/25 14:43		Jaswal	OK
35	Q3268-01DUP	WATER-TREATMENT	DUP	10/08/25 14:47		Jaswal	OK
36	Q3268-01L	WATER-TREATMENT	SD	10/08/25 14:52		Jaswal	OK
37	Q3268-01MS	WATER-TREATMENT	MS	10/08/25 14:56		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

38	Q3268-01MSD	WATER-TREATMENT	MSD	10/08/25 15:00		Jaswal	OK
39	CCV03	CCV03	CCV	10/08/25 15:04		Jaswal	OK
40	CCB03	CCB03	CCB	10/08/25 15:13		Jaswal	OK
41	Q3268-01A	WATER-TREATMENT	PS	10/08/25 15:17	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
42	PB169978BL	PB169978BL	MB	10/08/25 15:30		Jaswal	OK
43	PB169978BS	PB169978BS	LCS	10/08/25 15:34		Jaswal	OK
44	PB169959BL	PB169959BL	MB	10/08/25 15:38		Jaswal	OK
45	PB169959BS	PB169959BS	LCS	10/08/25 15:43		Jaswal	OK
46	Q3275-01	MW3	SAM	10/08/25 15:47		Jaswal	OK
47	Q3275-02	MW4	SAM	10/08/25 15:51		Jaswal	OK
48	Q3275-03	MW5	SAM	10/08/25 15:55	Na High	Jaswal	Dilution
49	CCV04	CCV04	CCV	10/08/25 16:11		Jaswal	OK
50	CCB04	CCB04	CCB	10/08/25 16:15		Jaswal	OK
51	Q3275-04	MW1	SAM	10/08/25 16:19	Not Use	Jaswal	Not Ok
52	Q3278-01	MW10	SAM	10/08/25 16:24	Not Use	Jaswal	Not Ok
53	Q3260-02	COMP	SAM	10/08/25 16:28		Jaswal	OK
54	Q3260-02DUP	COMP DUP	DUP	10/08/25 16:32		Jaswal	OK
55	Q3260-02L	COMPL	SD	10/08/25 16:37		Jaswal	OK
56	Q3260-02MS	COMPMS	MS	10/08/25 16:41		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137466

Review By	Janvi	Review On	10/10/2025 1:31:26 PM
Supervise By	jaswal	Supervise On	10/10/2025 10:40:09 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87038		
CCV Standard	MP87042		
ICSA Standard	MP87040,MP87048		
CRI Standard	MP87038		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

57	Q3260-02MSD	COMPMSD	MSD	10/08/25 16:45		Jaswal	OK
58	Q3260-02A	COMPA	PS	10/08/25 16:49	0.1ML M6008,M6017-10ML SAMPLE	Jaswal	OK
59	Q3267-01DL2	Q4DL2	SAM	10/08/25 16:53	Not Use	Jaswal	Not Ok
60	Q3267-01DL	Q4DL	SAM	10/08/25 16:58	50X Straight Dilution	Jaswal	OK
61	CCV05	CCV05	CCV	10/08/25 17:02		Jaswal	OK
62	CCB05	CCB05	CCB	10/08/25 17:06		Jaswal	OK
63	PB170024BL	PB170024BL	MB	10/08/25 17:10		Jaswal	OK
64	PB170024BS	PB170024BS	LCS	10/08/25 17:15		Jaswal	OK
65	Q3293-02	CONC-STORAGE-PA	SAM	10/08/25 17:19		Jaswal	OK
66	Q3297-10	TP1	SAM	10/08/25 17:23		Jaswal	OK
67	Q3297-20	TP2	SAM	10/08/25 17:27		Jaswal	OK
68	Q3302-06	OR-658-COMP-01	SAM	10/08/25 17:32	NOT USE	Jaswal	Not Ok
69	Q3302-06DUP	OR-658-COMP-01DU	DUP	10/08/25 17:36	NOT USE	Jaswal	Not Ok
70	Q3302-06L	OR-658-COMP-01L	SD	10/08/25 17:41	NOT USE	Jaswal	Not Ok
71	Q3302-06MS	OR-658-COMP-01MS	MS	10/08/25 17:45	NOT USE	Jaswal	Not Ok
72	Q3302-06MSD	OR-658-COMP-01MS	MSD	10/08/25 17:49	NOT USE	Jaswal	Not Ok
73	CCV06	CCV06	CCV	10/08/25 17:53		Jaswal	OK
74	CCB06	CCB06	CCB	10/08/25 17:57		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137484

Review By	jaswal	Review On	10/10/2025 1:52:59 PM
Supervise By	Janvi	Supervise On	10/13/2025 10:50:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87035		
CCV Standard	MP87042		
ICSA Standard	MP87041,MP87048		
CRI Standard	MP87035		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/09/25 11:10		Jaswal	OK
2	S1	S1	CAL2	10/09/25 11:15		Jaswal	OK
3	S2	S2	CAL3	10/09/25 11:19		Jaswal	OK
4	S3	S3	CAL4	10/09/25 11:23		Jaswal	OK
5	S4	S4	CAL5	10/09/25 11:27		Jaswal	OK
6	S5	S5	CAL6	10/09/25 11:31		Jaswal	OK
7	ICV01	ICV01	ICV	10/09/25 11:44		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/09/25 11:49		Jaswal	OK
9	ICB01	ICB01	ICB	10/09/25 11:59		Jaswal	OK
10	CRI01	CRI01	CRDL	10/09/25 12:04		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/09/25 12:09		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/09/25 12:19		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/09/25 12:23		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/09/25 12:30		Jaswal	OK
15	CCV01	CCV01	CCV	10/09/25 12:34		Jaswal	OK
16	CCB01	CCB01	CCB	10/09/25 12:46		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/09/25 12:56		Jaswal	OK
18	Q3278-01	MW10	SAM	10/09/25 13:05	Mn high	Jaswal	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137484

Review By	jaswal	Review On	10/10/2025 1:52:59 PM
Supervise By	Janvi	Supervise On	10/13/2025 10:50:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87035		
CCV Standard	MP87042		
ICSA Standard	MP87041,MP87048		
CRI Standard	MP87035		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

19	Q3275-04	MW1	SAM	10/09/25 13:09		Jaswal	OK
20	Q3278-01DL	MW10DL	SAM	10/09/25 13:14	2x for Mn	Jaswal	Confirms
21	Q3275-03DL	MW5DL	SAM	10/09/25 13:18	NOT USE	Jaswal	Not Ok
22	Q3302-06	OR-658-COMP-01	SAM	10/09/25 13:22		Jaswal	OK
23	Q3302-06DUP	OR-658-COMP-01DU	DUP	10/09/25 13:27		Jaswal	OK
24	Q3302-06L	OR-658-COMP-01L	SD	10/09/25 13:31		Jaswal	OK
25	LR-1	LR-1	HIGH STD	10/09/25 13:35		Jaswal	OK
26	CCV02	CCV02	CCV	10/09/25 14:16		Jaswal	OK
27	CCB02	CCB02	CCB	10/09/25 14:20		Jaswal	OK
28	Q3302-06MS	OR-658-COMP-01MS	MS	10/09/25 14:28		Jaswal	OK
29	Q3302-06MSD	OR-658-COMP-01MS	MSD	10/09/25 14:32		Jaswal	OK
30	Q3302-06A	OR-658-COMP-01A	PS	10/09/25 14:52	0.1ML M6008,M6017-10ML	Jaswal	OK
31	LR-3	LR-3	HIGH STD	10/09/25 14:58		Jaswal	OK
32	PB170010TB	PB170010TB	MB	10/09/25 15:05		Jaswal	OK
33	PB170023BL	PB170023BL	MB	10/09/25 15:10		Jaswal	OK
34	PB170023BS	PB170023BS	LCS	10/09/25 15:14		Jaswal	OK
35	PB170009BS	PB170009BS	LCS	10/09/25 15:18		Jaswal	OK
36	Q3302-01	OR-658-COMP-01	SAM	10/09/25 15:22		Jaswal	OK
37	CCV03	CCV03	CCV	10/09/25 15:26		Jaswal	OK
38	CCB03	CCB03	CCB	10/09/25 15:30		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137484

Review By	jaswal	Review On	10/10/2025 1:52:59 PM
Supervise By	Janvi	Supervise On	10/13/2025 10:50:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87035		
CCV Standard	MP87042		
ICSA Standard	MP87041,MP87048		
CRI Standard	MP87035		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

39	Q3303-01	HL-WEST	SAM	10/09/25 15:34		Jaswal	OK
40	Q3303-03	HL-EAST	SAM	10/09/25 15:38		Jaswal	OK
41	Q3304-01	TR-04-100725	SAM	10/09/25 15:42		Jaswal	OK
42	Q3304-01DUP	TR-04-100725DUP	DUP	10/09/25 15:47		Jaswal	OK
43	Q3304-01L	TR-04-100725L	SD	10/09/25 15:51		Jaswal	OK
44	Q3304-01MS	TR-04-100725MS	MS	10/09/25 15:55		Jaswal	OK
45	Q3304-01MSD	TR-04-100725MSD	MSD	10/09/25 15:59		Jaswal	OK
46	Q3304-01A	TR-04-100725A	PS	10/09/25 16:02	0.1ML M6008,M6017-10ML	Jaswal	OK
47	Q3304-03	TR-05-100725	SAM	10/09/25 16:06		Jaswal	OK
48	Q3305-01	AU-06-100725	SAM	10/09/25 16:10		Jaswal	OK
49	CCV04	CCV04	CCV	10/09/25 16:26		Jaswal	OK
50	CCB04	CCB04	CCB	10/09/25 16:41		Jaswal	OK
51	Q3292-01	RW8-SP100-2025100	SAM	10/09/25 16:45		Jaswal	OK
52	Q3292-02	RW8-SP303-2025100	SAM	10/09/25 16:50		Jaswal	OK
53	Q3292-02DUP	RW8-SP303-2025100	DUP	10/09/25 16:54		Jaswal	OK
54	Q3292-02L	RW8-SP303-2025100	SD	10/09/25 16:58		Jaswal	OK
55	Q3292-02MS	RW8-SP303-2025100	MS	10/09/25 17:02		Jaswal	OK
56	Q3292-02MSD	RW8-SP303-2025100	MSD	10/09/25 17:06		Jaswal	OK
57	Q3292-02A	RW8-SP303-2025100	PS	10/09/25 17:10	0.1ML M6008,M6017-10ML	Jaswal	OK
58	Q3275-03DL2	MW5DL2	SAM	10/09/25 17:14	10x for Na	Jaswal	Confirms

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137484

Review By	jaswal	Review On	10/10/2025 1:52:59 PM
Supervise By	Janvi	Supervise On	10/13/2025 10:50:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87031,MP87038,MP87035,MP87034,MP87033,MP87032		
ICV Standard	MP87047,MP87035		
CCV Standard	MP87042		
ICSA Standard	MP87041,MP87048		
CRI Standard	MP87035		
LCS Standard			
Chk Standard	MP87043,MP87044,MP87045,MP87046,MP87049		

59	PB170044BS	PB170044BS	LCS	10/09/25 17:23		Jaswal	OK
60	CCV05	CCV05	CCV	10/09/25 17:27		Jaswal	OK
61	CCB05	CCB05	CCB	10/09/25 17:31		Jaswal	OK
62	Q3313-01	ETGI-360	SAM	10/09/25 17:35		Jaswal	OK
63	Q3315-03	72-11989	SAM	10/09/25 17:39		Jaswal	OK
64	Q3315-03DUP	72-11989DUP	DUP	10/09/25 17:43		Jaswal	OK
65	Q3315-03L	72-11989L	SD	10/09/25 17:48		Jaswal	OK
66	Q3315-03MS	72-11989MS	MS	10/09/25 17:52		Jaswal	OK
67	Q3315-03MSD	72-11989MSD	MSD	10/09/25 17:56		Jaswal	OK
68	Q3315-03A	72-11989A	PS	10/09/25 17:59	0.1ML M6008,M6017-10ML	Jaswal	OK
69	Q3318-01	VNJ-207	SAM	10/09/25 18:03		Jaswal	OK
70	Q3319-01	OK-01-100825	SAM	10/09/25 18:07		Jaswal	OK
71	PB170009BL	PB170009BL	MB	10/09/25 18:11		Jaswal	OK
72	CCV06	CCV06	CCV	10/09/25 18:16		Jaswal	OK
73	CCB06	CCB06	CCB	10/09/25 18:20		Jaswal	OK
74	PB170044BL	PB170044BL	MB	10/09/25 18:24		Jaswal	OK
75	CCV07	CCV07	CCV	10/09/25 18:28		Jaswal	OK
76	CCB07	CCB07	CCB	10/09/25 18:33		Jaswal	OK

SOP ID : M3010A-Digestion-17

SDG No : N/A

Matrix : WATER

Pipette ID: ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID: 1. HOT BOCK #1 2. N/A

Start Digest Date: 10/03/2025 Time : 10:20 Temp : 96 °C

End Digest Date: 10/03/2025 Time : 13:25 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: SKJ

Supervisor Signature: [Signature]

Temp : 1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6158
1:1 HCL	5.00	MP87148
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96 C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/03/2025 13:25	SKJ met dig	SKJ (metals lab)
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB169978BL	PBW978	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	1
PB169978BS	LCS978	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	2
Q3268-01MS	WATER-TREATMENT-DISCHA RCEMS	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	5
Q3268-01MSD	WATER-TREATMENT-DISCHA RCEMSD	<2	50	25	Colorless	Colorless	Clear	Clear	M6180,M6181	6
Q3268-01DUP	WATER-TREATMENT-DISCHA RCE DUP	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	4
Q3268-01	WATER-TREATMENT-DISCHA RCE	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	3
Q3275-01	MW3	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	7
Q3275-02	MW4	<2	50	25	Brown	light Brown	Cloudy	Clear	N/A	8
Q3275-03	MW5	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	9
Q3275-04	MW1	<2	50	25	light Brown	Colorless	Clear	Clear	N/A	10
Q3278-01	MW10	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	11