

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19000	20000	95	90 - 110	P	10/08/2025	12:24	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1770	2000	89	80 - 120	P	10/08/2025	12:32	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	25900	25000	104	90 - 110	P	10/08/2025	13:12	LB137466
CCV02	Sodium	24100	25000	96	90 - 110	P	10/08/2025	14:14	LB137466
CCV03	Sodium	22600	25000	90	90 - 110	P	10/08/2025	15:04	LB137466
CCV04	Sodium	26100	25000	104	90 - 110	P	10/08/2025	16:11	LB137466
CCV05	Sodium	26900	25000	108	90 - 110	P	10/08/2025	17:02	LB137466
CCV06	Sodium	26000	25000	104	90 - 110	P	10/08/2025	17:53	LB137466

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Sodium	19600	20000	98	90 - 110	P	10/09/2025	11:44	LB137484

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Sodium	1610	2000	80	80 - 120	P	10/09/2025	11:49	LB137484

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3275

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Sodium	24400	25000	98	90 - 110	P	10/09/2025	12:34	LB137484
CCV02	Sodium	23500	25000	94	90 - 110	P	10/09/2025	14:16	LB137484
CCV03	Sodium	23400	25000	94	90 - 110	P	10/09/2025	15:26	LB137484
CCV04	Sodium	24300	25000	97	90 - 110	P	10/09/2025	16:26	LB137484
CCV05	Sodium	23100	25000	92	90 - 110	P	10/09/2025	17:27	LB137484
CCV06	Sodium	26600	25000	106	90 - 110	P	10/09/2025	18:16	LB137484
CCV07	Sodium	26300	25000	105	90 - 110	P	10/09/2025	18:28	LB137484