

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3276</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/14/2025</u> <u>10:02</u>
Lab File ID:	<u>VX048153.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.667		28.76	20
Chloromethane	0.680	0.757	0.1	11.32	20
Vinyl Chloride	0.666	0.847		27.18	20
Bromomethane	0.422	0.520		23.22	20
Chloroethane	0.445	0.518		16.4	20
Trichlorofluoromethane	0.989	1.294		30.84	20
1,1,2-Trichlorotrifluoroethane	0.578	0.719		24.39	20
1,1-Dichloroethene	0.594	0.702		18.18	20
Acetone	0.346	0.318		-8.09	20
Carbon Disulfide	1.626	2.086		28.29	20
Methyl tert-butyl Ether	2.169	2.175		0.28	20
Methyl Acetate	0.770	0.685		-11.04	20
Methylene Chloride	0.732	0.779		6.42	20
trans-1,2-Dichloroethene	0.640	0.738		15.31	20
1,1-Dichloroethane	1.263	1.365	0.1	8.08	20
Cyclohexane	1.052	1.169		11.12	20
2-Butanone	0.432	0.397		-8.1	20
Carbon Tetrachloride	0.532	0.645		21.24	20
cis-1,2-Dichloroethene	0.783	0.834		6.51	20
Bromochloromethane	0.551	0.528		-4.17	20
Chloroform	1.276	1.395		9.33	20
1,1,1-Trichloroethane	1.082	1.227		13.4	20
Methylcyclohexane	0.556	0.660		18.7	20
Benzene	1.488	1.679		12.84	20
1,2-Dichloroethane	0.566	0.618		9.19	20
Trichloroethene	0.360	0.415		15.28	20
1,2-Dichloropropane	0.381	0.420		10.24	20
Bromodichloromethane	0.572	0.658		15.03	20
4-Methyl-2-Pentanone	0.477	0.470		-1.47	20
Toluene	0.917	1.028		12.1	20
t-1,3-Dichloropropene	0.584	0.626		7.19	20
cis-1,3-Dichloropropene	0.624	0.683		9.45	20
1,1,2-Trichloroethane	0.354	0.381		7.63	20
2-Hexanone	0.343	0.327		-4.66	20
Dibromochloromethane	0.407	0.473		16.22	20
1,2-Dibromoethane	0.360	0.394		9.44	20
Tetrachloroethene	0.326	0.383		17.49	20
Chlorobenzene	1.136	1.236	0.3	8.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>GENV01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3276</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/14/2025</u> <u>10:02</u>
Lab File ID:	<u>VX048153.D</u>	Init. Calib. Date(s):	<u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23</u> <u>12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.162		10.31	20
m/p-Xylenes	0.729	0.814		11.66	20
o-Xylene	0.706	0.754		6.8	20
Styrene	1.236	1.320		6.8	20
Bromoform	0.293	0.316	0.1	7.85	20
Isopropylbenzene	3.801	3.991		5	20
1,1,2,2-Tetrachloroethane	1.150	1.134	0.3	-1.39	20
1,3-Dichlorobenzene	1.739	1.779		2.3	20
1,4-Dichlorobenzene	1.785	1.799		0.78	20
1,2-Dichlorobenzene	1.657	1.679		1.33	20
1,2-Dibromo-3-Chloropropane	0.233	0.210		-9.87	20
1,2,4-Trichlorobenzene	1.077	1.027		-4.64	20
1,2,3-Trichlorobenzene	1.002	0.966		-3.59	20
1,2-Dichloroethane-d4	0.796	0.845		6.16	20
Dibromofluoromethane	0.335	0.391		16.72	20
Toluene-d8	1.160	1.287		10.95	20
4-Bromofluorobenzene	0.440	0.491		11.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.