

Surrogate Summary

SDG No.: **Q3276**

Client: **G Environmental**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3276-01	SB2	1,2-Dichloroethane-d4	50	52.1	104		63	155
		Dibromofluoromethane	50	51.1	102		70	134
		Toluene-d8	50	53.9	108		74	123
		4-Bromofluorobenzene	50	57.8	116		17	146
Q3276-02	SB2A	1,2-Dichloroethane-d4	50	57.5	115		63	155
		Dibromofluoromethane	50	51.7	103		70	134
		Toluene-d8	50	51.9	104		74	123
		4-Bromofluorobenzene	50	61.6	123		17	146
VX1014MBL01	VX1014MBL01	1,2-Dichloroethane-d4	50	56.7	113		63	155
		Dibromofluoromethane	50	55.6	111		70	134
		Toluene-d8	50	53.8	108		74	123
		4-Bromofluorobenzene	50	59.0	118		17	146
VX1014MBS02	VX1014MBS02	1,2-Dichloroethane-d4	50	45.8	92		63	155
		Dibromofluoromethane	50	49.6	99		70	134
		Toluene-d8	50	48.0	96		74	123
		4-Bromofluorobenzene	50	49.6	99		17	146
VY1014SBL01	VY1014SBL01	1,2-Dichloroethane-d4	50	59.3	119		63	155
		Dibromofluoromethane	50	54.2	108		70	134
		Toluene-d8	50	51.9	104		74	123
		4-Bromofluorobenzene	50	52.1	104		17	146
VY1014SBS01	VY1014SBS01	1,2-Dichloroethane-d4	50	46.2	92		63	155
		Dibromofluoromethane	50	48.9	98		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	46.5	93		17	146
VY1014SBSD01	VY1014SBSD01	1,2-Dichloroethane-d4	50	46.2	92		63	155
		Dibromofluoromethane	50	49.1	98		70	134
		Toluene-d8	50	49.3	99		74	123
		4-Bromofluorobenzene	50	46.7	93		17	146