

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3276
 Lab File ID: VX048153.D Date Analyzed: 10/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	124958	5.53	212282	6.74	195629	10.04
UPPER LIMIT	249916	6.031	424564	7.239	391258	10.537
LOWER LIMIT	62479	5.031	106141	6.239	97814.5	9.537
EPA SAMPLE NO.						
SB2A	111354	5.54	240688	6.75	245501	10.04
VX1014MBL01	92850	5.53	189331	6.75	199229	10.04
VX1014MBS02	126511	5.54	219762	6.75	203522	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.