

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3276
 Lab File ID: VY023456.D Date Analyzed: 10/14/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	571722	7.71	808604	8.62	715207	11.42
UPPER LIMIT	1143440	8.213	1617210	9.116	1430410	11.92
LOWER LIMIT	285861	7.213	404302	8.116	357604	10.92
EPA SAMPLE NO.						
SB2	582031	7.71	1136657	8.62	1183075	11.42
VY1014SBL01	564646	7.71	1109577	8.62	1120800	11.42
VY1014SBS01	531866	7.71	775568	8.62	681796	11.42
VY1014SBSD01	517858	7.71	759118	8.62	662146	11.42

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.