

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3276
 Lab File ID: VX048153.D Date Analyzed: 10/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	95000	12.00€				
UPPER LIMIT	190000	12.50€				
LOWER LIMIT	47500	11.50€				
EPA SAMPLE NO.						
SB2A	125319	12.01				
VX1014MBL01	93640	12.01				
VX1014MBS02	105658	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.