

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3276
 Lab File ID: VY023456.D Date Analyzed: 10/14/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:25
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	372874	13.346				
UPPER LIMIT	745748	13.846				
LOWER LIMIT	186437	12.846				
EPA SAMPLE NO.						
SB2	559850	13.35				
VY1014SBL01	479501	13.35				
VY1014SBS01	357890	13.35				
VY1014SBSD01	344738	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.