

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3277

Project: Newport

RunNo.: LB137425

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.02	7	100	90-110	10/03/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	10/03/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	10/03/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3277

Project: Newport

RunNo.: LB137430

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	10/06/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/06/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	10/06/2025

Initial and Continuing Calibration Verification

Client: G Environmental

SDG No.: Q3277

Project: Newport

RunNo.: LB137474

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.01	7	100	90-110	10/09/2025
Sample ID:	CCV1						
pH		pH	2.02	2.00	101	90-110	10/09/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	10/09/2025