

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3277
Project:	Newport	Sample ID:	Q3256-04
Client ID:	SOIL-PILE-1-WCDUP	Percent Solids for Spike Sample:	100

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Reactive Sulfide	mg/Kg	+/-20	3.16	J	3.16	J	1	0		10/03/2025

Duplicate Sample Summary

Client:	G Environmental	SDG No.:	Q3277
Project:	Newport	Sample ID:	Q3269-01
Client ID:	EG1-TP1DUP	Percent Solids for Spike Sample:	90.8

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ignitability	oC	+/-20	NO		NO		1	0		10/06/2025

Duplicate Sample Summary

Client: G Environmental Project: Newport Client ID: RT5417-SOILDUP	SDG No.: Q3277 Sample ID: Q3272-05 Percent Solids for Spike Sample: 100
---	--

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Corrosivity	pH	+/-20	12.5		12.5		1	0.24		10/03/2025
Reactive Cyanide	mg/Kg	+/-20	0.015	J	0.015	J	1	0		10/06/2025

Duplicate Sample Summary

Client: G Environmental Project: Newport Client ID: WC1DUP	SDG No.: Q3277 Sample ID: Q3277-01 Percent Solids for Spike Sample: 97.6
---	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Paint Filter	ml/100gm	+/-20	1.00	U	1.00	U	1	0		10/03/2025
pH	pH	+/-20	7.22		7.24		1	0.28		10/09/2025