

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV59	Mercury	4.07	4.0	102	90 - 110	CV	10/06/2025	14:56	LB137429

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV83	Mercury	4.89	5.0	98	90 - 110	CV	10/06/2025	15:01	LB137429
CCV84	Mercury	4.84	5.0	97	90 - 110	CV	10/06/2025	15:30	LB137429
CCV85	Mercury	4.83	5.0	97	90 - 110	CV	10/06/2025	15:59	LB137429

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7920	8000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Antimony	4110	4000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Arsenic	4000	4000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Barium	7940	8000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Beryllium	201	200	100	90 - 110	P	10/07/2025	10:54	LB137447
	Cadmium	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Calcium	19900	20000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Chromium	817	800	102	90 - 110	P	10/07/2025	10:54	LB137447
	Cobalt	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Copper	1020	1000	102	90 - 110	P	10/07/2025	10:54	LB137447
	Iron	4130	4000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Lead	3890	4000	97	90 - 110	P	10/07/2025	10:54	LB137447
	Magnesium	19900	20000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Manganese	1990	2000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Nickel	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Potassium	20200	20000	101	90 - 110	P	10/07/2025	10:54	LB137447
	Selenium	4000	4000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Silver	1030	1000	103	90 - 110	P	10/07/2025	10:54	LB137447
	Sodium	20100	20000	100	90 - 110	P	10/07/2025	10:54	LB137447
	Thallium	3720	4000	93	90 - 110	P	10/07/2025	10:54	LB137447
	Vanadium	1980	2000	99	90 - 110	P	10/07/2025	10:54	LB137447
	Zinc	2020	2000	101	90 - 110	P	10/07/2025	10:54	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	92.2	100	92	80 - 120	P	10/07/2025	12:01	LB137447
	Antimony	49.4	50.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Arsenic	18.4	20.0	92	80 - 120	P	10/07/2025	12:01	LB137447
	Barium	95.7	100	96	80 - 120	P	10/07/2025	12:01	LB137447
	Beryllium	6.20	6.0	103	80 - 120	P	10/07/2025	12:01	LB137447
	Cadmium	5.85	6.0	98	80 - 120	P	10/07/2025	12:01	LB137447
	Calcium	2020	2000	101	80 - 120	P	10/07/2025	12:01	LB137447
	Chromium	9.65	10.0	96	80 - 120	P	10/07/2025	12:01	LB137447
	Cobalt	29.0	30.0	97	80 - 120	P	10/07/2025	12:01	LB137447
	Copper	21.9	20.0	110	80 - 120	P	10/07/2025	12:01	LB137447
	Iron	119	100	119	80 - 120	P	10/07/2025	12:01	LB137447
	Lead	11.1	12.0	93	80 - 120	P	10/07/2025	12:01	LB137447
	Magnesium	2070	2000	104	80 - 120	P	10/07/2025	12:01	LB137447
	Manganese	20.8	20.0	104	80 - 120	P	10/07/2025	12:01	LB137447
	Nickel	39.6	40.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Potassium	1910	2000	95	80 - 120	P	10/07/2025	12:01	LB137447
	Selenium	23.6	20.0	118	80 - 120	P	10/07/2025	12:01	LB137447
	Silver	9.88	10.0	99	80 - 120	P	10/07/2025	12:01	LB137447
	Sodium	1890	2000	94	80 - 120	P	10/07/2025	12:01	LB137447
	Thallium	43.0	40.0	107	80 - 120	P	10/07/2025	12:01	LB137447
	Vanadium	37.3	40.0	93	80 - 120	P	10/07/2025	12:01	LB137447
	Zinc	36.9	40.0	92	80 - 120	P	10/07/2025	12:01	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Antimony	5050	5000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Arsenic	5080	5000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Barium	9980	10000	100	90 - 110	P	10/07/2025	12:50	LB137447
	Beryllium	252	250	101	90 - 110	P	10/07/2025	12:50	LB137447
	Cadmium	2510	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Calcium	25200	25000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	12:50	LB137447
	Cobalt	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Copper	1270	1250	101	90 - 110	P	10/07/2025	12:50	LB137447
	Iron	5200	5000	104	90 - 110	P	10/07/2025	12:50	LB137447
	Lead	4980	5000	100	90 - 110	P	10/07/2025	12:50	LB137447
	Magnesium	25200	25000	101	90 - 110	P	10/07/2025	12:50	LB137447
	Manganese	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Nickel	2490	2500	100	90 - 110	P	10/07/2025	12:50	LB137447
	Potassium	25500	25000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Selenium	5150	5000	103	90 - 110	P	10/07/2025	12:50	LB137447
	Silver	1280	1250	103	90 - 110	P	10/07/2025	12:50	LB137447
	Sodium	25600	25000	102	90 - 110	P	10/07/2025	12:50	LB137447
	Thallium	4870	5000	97	90 - 110	P	10/07/2025	12:50	LB137447
	Vanadium	2520	2500	101	90 - 110	P	10/07/2025	12:50	LB137447
	Zinc	2600	2500	104	90 - 110	P	10/07/2025	12:50	LB137447
CCV02	Aluminum	9930	10000	99	90 - 110	P	10/07/2025	14:19	LB137447
	Antimony	5050	5000	101	90 - 110	P	10/07/2025	14:19	LB137447
	Arsenic	5080	5000	102	90 - 110	P	10/07/2025	14:19	LB137447
	Barium	9770	10000	98	90 - 110	P	10/07/2025	14:19	LB137447
	Beryllium	250	250	100	90 - 110	P	10/07/2025	14:19	LB137447
	Cadmium	2510	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Calcium	24900	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	14:19	LB137447
	Cobalt	2480	2500	99	90 - 110	P	10/07/2025	14:19	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	14:19	LB137447
	Iron	5340	5000	107	90 - 110	P	10/07/2025	14:19	LB137447
	Lead	4960	5000	99	90 - 110	P	10/07/2025	14:19	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	24900	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Manganese	2480	2500	99	90 - 110	P	10/07/2025	14:19	LB137447
	Nickel	2490	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Potassium	26100	25000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Selenium	5190	5000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Silver	1290	1250	103	90 - 110	P	10/07/2025	14:19	LB137447
	Sodium	25000	25000	100	90 - 110	P	10/07/2025	14:19	LB137447
	Thallium	5200	5000	104	90 - 110	P	10/07/2025	14:19	LB137447
	Vanadium	2500	2500	100	90 - 110	P	10/07/2025	14:19	LB137447
	Zinc	2600	2500	104	90 - 110	P	10/07/2025	14:19	LB137447
CCV03	Aluminum	10000	10000	100	90 - 110	P	10/07/2025	15:08	LB137447
	Antimony	4960	5000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Arsenic	4970	5000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Barium	9860	10000	99	90 - 110	P	10/07/2025	15:08	LB137447
	Beryllium	253	250	101	90 - 110	P	10/07/2025	15:08	LB137447
	Cadmium	2440	2500	98	90 - 110	P	10/07/2025	15:08	LB137447
	Calcium	25100	25000	100	90 - 110	P	10/07/2025	15:08	LB137447
	Chromium	1010	1000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Cobalt	2420	2500	97	90 - 110	P	10/07/2025	15:08	LB137447
	Copper	1240	1250	99	90 - 110	P	10/07/2025	15:08	LB137447
	Iron	5120	5000	102	90 - 110	P	10/07/2025	15:08	LB137447
	Lead	4840	5000	97	90 - 110	P	10/07/2025	15:08	LB137447
	Magnesium	25100	25000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Manganese	2510	2500	100	90 - 110	P	10/07/2025	15:08	LB137447
	Nickel	2430	2500	97	90 - 110	P	10/07/2025	15:08	LB137447
	Potassium	24200	25000	97	90 - 110	P	10/07/2025	15:08	LB137447
	Selenium	5070	5000	101	90 - 110	P	10/07/2025	15:08	LB137447
	Silver	1250	1250	100	90 - 110	P	10/07/2025	15:08	LB137447
	Sodium	23700	25000	95	90 - 110	P	10/07/2025	15:08	LB137447
	Thallium	4850	5000	97	90 - 110	P	10/07/2025	15:08	LB137447
	Vanadium	2540	2500	102	90 - 110	P	10/07/2025	15:08	LB137447
	Zinc	2520	2500	101	90 - 110	P	10/07/2025	15:08	LB137447
CCV04	Aluminum	9980	10000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Antimony	5080	5000	102	90 - 110	P	10/07/2025	16:00	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5090	5000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Barium	9920	10000	99	90 - 110	P	10/07/2025	16:00	LB137447
	Beryllium	254	250	102	90 - 110	P	10/07/2025	16:00	LB137447
	Cadmium	2490	2500	100	90 - 110	P	10/07/2025	16:00	LB137447
	Calcium	25100	25000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Chromium	1020	1000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Cobalt	2480	2500	99	90 - 110	P	10/07/2025	16:00	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	16:00	LB137447
	Iron	5080	5000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Lead	4950	5000	99	90 - 110	P	10/07/2025	16:00	LB137447
	Magnesium	25200	25000	101	90 - 110	P	10/07/2025	16:00	LB137447
	Manganese	2530	2500	101	90 - 110	P	10/07/2025	16:00	LB137447
	Nickel	2480	2500	99	90 - 110	P	10/07/2025	16:00	LB137447
	Potassium	24300	25000	97	90 - 110	P	10/07/2025	16:00	LB137447
	Selenium	5180	5000	104	90 - 110	P	10/07/2025	16:00	LB137447
	Silver	1260	1250	101	90 - 110	P	10/07/2025	16:00	LB137447
	Sodium	23700	25000	95	90 - 110	P	10/07/2025	16:00	LB137447
	Thallium	5000	5000	100	90 - 110	P	10/07/2025	16:00	LB137447
	Vanadium	2540	2500	102	90 - 110	P	10/07/2025	16:00	LB137447
	Zinc	2550	2500	102	90 - 110	P	10/07/2025	16:00	LB137447
CCV05	Aluminum	9820	10000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Antimony	4960	5000	99	90 - 110	P	10/07/2025	16:50	LB137447
	Arsenic	5000	5000	100	90 - 110	P	10/07/2025	16:50	LB137447
	Barium	9540	10000	95	90 - 110	P	10/07/2025	16:50	LB137447
	Beryllium	245	250	98	90 - 110	P	10/07/2025	16:50	LB137447
	Cadmium	2480	2500	99	90 - 110	P	10/07/2025	16:50	LB137447
	Calcium	24500	25000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Chromium	1030	1000	103	90 - 110	P	10/07/2025	16:50	LB137447
	Cobalt	2450	2500	98	90 - 110	P	10/07/2025	16:50	LB137447
	Copper	1250	1250	100	90 - 110	P	10/07/2025	16:50	LB137447
	Iron	5320	5000	106	90 - 110	P	10/07/2025	16:50	LB137447
	Lead	4890	5000	98	90 - 110	P	10/07/2025	16:50	LB137447
	Magnesium	24700	25000	99	90 - 110	P	10/07/2025	16:50	LB137447
	Manganese	2430	2500	97	90 - 110	P	10/07/2025	16:50	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2460	2500	98	90 - 110	P	10/07/2025	16:50	LB137447
	Potassium	26200	25000	105	90 - 110	P	10/07/2025	16:50	LB137447
	Selenium	5090	5000	102	90 - 110	P	10/07/2025	16:50	LB137447
	Silver	1280	1250	103	90 - 110	P	10/07/2025	16:50	LB137447
	Sodium	26000	25000	104	90 - 110	P	10/07/2025	16:50	LB137447
	Thallium	5150	5000	103	90 - 110	P	10/07/2025	16:50	LB137447
	Vanadium	2490	2500	99	90 - 110	P	10/07/2025	16:50	LB137447
	Zinc	2620	2500	105	90 - 110	P	10/07/2025	16:50	LB137447
CCV06	Aluminum	9870	10000	99	90 - 110	P	10/07/2025	17:49	LB137447
	Antimony	5030	5000	100	90 - 110	P	10/07/2025	17:49	LB137447
	Arsenic	5040	5000	101	90 - 110	P	10/07/2025	17:49	LB137447
	Barium	9460	10000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Beryllium	244	250	98	90 - 110	P	10/07/2025	17:49	LB137447
	Cadmium	2460	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Calcium	24400	25000	98	90 - 110	P	10/07/2025	17:49	LB137447
	Chromium	1000	1000	100	90 - 110	P	10/07/2025	17:49	LB137447
	Cobalt	2440	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Copper	1250	1250	100	90 - 110	P	10/07/2025	17:49	LB137447
	Iron	4840	5000	97	90 - 110	P	10/07/2025	17:49	LB137447
	Lead	4880	5000	98	90 - 110	P	10/07/2025	17:49	LB137447
	Magnesium	24700	25000	99	90 - 110	P	10/07/2025	17:49	LB137447
	Manganese	2420	2500	97	90 - 110	P	10/07/2025	17:49	LB137447
	Nickel	2450	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Potassium	25200	25000	101	90 - 110	P	10/07/2025	17:49	LB137447
	Selenium	5150	5000	103	90 - 110	P	10/07/2025	17:49	LB137447
	Silver	1230	1250	98	90 - 110	P	10/07/2025	17:49	LB137447
	Sodium	23800	25000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Thallium	4750	5000	95	90 - 110	P	10/07/2025	17:49	LB137447
	Vanadium	2460	2500	98	90 - 110	P	10/07/2025	17:49	LB137447
	Zinc	2480	2500	99	90 - 110	P	10/07/2025	17:49	LB137447
CCV07	Aluminum	9840	10000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Antimony	5100	5000	102	90 - 110	P	10/07/2025	18:44	LB137447
	Arsenic	5120	5000	102	90 - 110	P	10/07/2025	18:44	LB137447
	Barium	9570	10000	96	90 - 110	P	10/07/2025	18:44	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	244	250	98	90 - 110	P	10/07/2025	18:44	LB137447
	Cadmium	2470	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Calcium	24400	25000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Chromium	997	1000	100	90 - 110	P	10/07/2025	18:44	LB137447
	Cobalt	2460	2500	98	90 - 110	P	10/07/2025	18:44	LB137447
	Copper	1260	1250	101	90 - 110	P	10/07/2025	18:44	LB137447
	Iron	4950	5000	99	90 - 110	P	10/07/2025	18:44	LB137447
	Lead	4920	5000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Magnesium	24600	25000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Manganese	2440	2500	98	90 - 110	P	10/07/2025	18:44	LB137447
	Nickel	2470	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Potassium	24000	25000	96	90 - 110	P	10/07/2025	18:44	LB137447
	Selenium	5280	5000	106	90 - 110	P	10/07/2025	18:44	LB137447
	Silver	1230	1250	98	90 - 110	P	10/07/2025	18:44	LB137447
	Sodium	24600	25000	98	90 - 110	P	10/07/2025	18:44	LB137447
	Thallium	5050	5000	101	90 - 110	P	10/07/2025	18:44	LB137447
	Vanadium	2480	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
	Zinc	2480	2500	99	90 - 110	P	10/07/2025	18:44	LB137447
CCV08	Aluminum	9510	10000	95	90 - 110	P	10/07/2025	19:20	LB137447
	Antimony	4850	5000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Arsenic	4850	5000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Barium	9160	10000	92	90 - 110	P	10/07/2025	19:20	LB137447
	Beryllium	238	250	95	90 - 110	P	10/07/2025	19:20	LB137447
	Cadmium	2370	2500	95	90 - 110	P	10/07/2025	19:20	LB137447
	Calcium	23500	25000	94	90 - 110	P	10/07/2025	19:20	LB137447
	Chromium	967	1000	97	90 - 110	P	10/07/2025	19:20	LB137447
	Cobalt	2360	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Copper	1210	1250	96	90 - 110	P	10/07/2025	19:20	LB137447
	Iron	4810	5000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Lead	4720	5000	94	90 - 110	P	10/07/2025	19:20	LB137447
	Magnesium	23900	25000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Manganese	2350	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Nickel	2360	2500	94	90 - 110	P	10/07/2025	19:20	LB137447
	Potassium	23100	25000	92	90 - 110	P	10/07/2025	19:20	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	4950	5000	99	90 - 110	P	10/07/2025	19:20	LB137447
	Silver	1200	1250	96	90 - 110	P	10/07/2025	19:20	LB137447
	Sodium	24000	25000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Thallium	4540	5000	91	90 - 110	P	10/07/2025	19:20	LB137447
	Vanadium	2380	2500	95	90 - 110	P	10/07/2025	19:20	LB137447
	Zinc	2420	2500	97	90 - 110	P	10/07/2025	19:20	LB137447