



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.20	0.2	98	70 - 130	CV	10/06/2025	15:06	LB137429
CRI01	Aluminum	90.0	100	90	65 - 135	P	10/07/2025	12:10	LB137447
	Antimony	49.9	50.0	100	65 - 135	P	10/07/2025	12:10	LB137447
	Arsenic	17.7	20.0	88	65 - 135	P	10/07/2025	12:10	LB137447
	Barium	97.4	100	97	65 - 135	P	10/07/2025	12:10	LB137447
	Beryllium	6.27	6.0	104	65 - 135	P	10/07/2025	12:10	LB137447
	Cadmium	5.96	6.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Calcium	2040	2000	102	65 - 135	P	10/07/2025	12:10	LB137447
	Chromium	9.90	10.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Cobalt	29.6	30.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Copper	22.2	20.0	111	65 - 135	P	10/07/2025	12:10	LB137447
	Iron	123	100	123	65 - 135	P	10/07/2025	12:10	LB137447
	Lead	12.1	12.0	101	65 - 135	P	10/07/2025	12:10	LB137447
	Magnesium	2090	2000	104	65 - 135	P	10/07/2025	12:10	LB137447
	Manganese	20.9	20.0	105	65 - 135	P	10/07/2025	12:10	LB137447
	Nickel	40.1	40.0	100	65 - 135	P	10/07/2025	12:10	LB137447
	Potassium	1880	2000	94	65 - 135	P	10/07/2025	12:10	LB137447
	Selenium	22.6	20.0	113	65 - 135	P	10/07/2025	12:10	LB137447
	Silver	9.85	10.0	98	65 - 135	P	10/07/2025	12:10	LB137447
	Sodium	1910	2000	95	65 - 135	P	10/07/2025	12:10	LB137447
	Thallium	45.0	40.0	113	65 - 135	P	10/07/2025	12:10	LB137447
	Vanadium	37.6	40.0	94	65 - 135	P	10/07/2025	12:10	LB137447
	Zinc	37.9	40.0	95	65 - 135	P	10/07/2025	12:10	LB137447