

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169979BS							
Aluminum	mg/Kg	100	97.8		98	80 - 120	P
Antimony	mg/Kg	40.0	41.0		102	80 - 120	P
Arsenic	mg/Kg	40.0	39.6		99	80 - 120	P
Barium	mg/Kg	10.0	9.09		91	80 - 120	P
Beryllium	mg/Kg	10.0	10.4		104	80 - 120	P
Cadmium	mg/Kg	10.0	9.53		95	80 - 120	P
Calcium	mg/Kg	50.0	50.3	J	101	80 - 120	P
Chromium	mg/Kg	20.0	22.1		110	80 - 120	P
Cobalt	mg/Kg	10.0	9.66		97	80 - 120	P
Copper	mg/Kg	15.0	15.5		103	80 - 120	P
Iron	mg/Kg	150	157		105	80 - 120	P
Lead	mg/Kg	50.0	47.4		95	80 - 120	P
Magnesium	mg/Kg	100	99.6	J	100	80 - 120	P
Manganese	mg/Kg	10.0	10.3		103	80 - 120	P
Nickel	mg/Kg	25.0	25.2		101	80 - 120	P
Potassium	mg/Kg	500	465		93	80 - 120	P
Selenium	mg/Kg	100	102		102	80 - 120	P
Silver	mg/Kg	3.8	3.78		100	80 - 120	P
Sodium	mg/Kg	150	131		87	80 - 120	P
Thallium	mg/Kg	100	84.5		84	80 - 120	P
Vanadium	mg/Kg	15.0	15.0		100	80 - 120	P
Zinc	mg/Kg	10.0	10.3		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169993BS Mercury	mg/Kg	0.27	0.25		91	80 - 120	CV