



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.18	0.2	92	70 - 130	CV	10/07/2025	10:13	LB137441
CRI01	Arsenic	17.7	20.0	88	65 - 135	P	10/07/2025	12:10	LB137447
	Barium	97.4	100	97	65 - 135	P	10/07/2025	12:10	LB137447
	Cadmium	5.96	6.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Chromium	9.90	10.0	99	65 - 135	P	10/07/2025	12:10	LB137447
	Lead	12.1	12.0	101	65 - 135	P	10/07/2025	12:10	LB137447
	Selenium	22.6	20.0	113	65 - 135	P	10/07/2025	12:10	LB137447
	Silver	9.85	10.0	98	65 - 135	P	10/07/2025	12:10	LB137447