



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB61	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:06	LB137441

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB90	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:11	LB137441
CCB91	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:47	LB137441
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:16	LB137441
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:39	LB137441
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:57	LB137441

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:06	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:06	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:06	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:06	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:06	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:06	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	12:58	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	12:58	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	12:58	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	12:58	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	12:58	LB137447
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	14:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	14:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	14:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	14:24	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	14:24	LB137447
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	15:12	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	15:12	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	15:12	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	15:12	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	15:12	LB137447
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	16:04	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	16:04	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	16:04	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	16:04	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	16:04	LB137447
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:01	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:01	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:01	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:01	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:01	LB137447
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	17:59	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	17:59	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	17:59	LB137447

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/07/2025	17:59	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	17:59	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	17:59	LB137447
CCB07	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	18:48	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	18:48	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	18:48	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	18:48	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	18:48	LB137447
CCB08	Arsenic	5.12	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Barium	14.6	+/-50	U	100	P	10/07/2025	19:24	LB137447
	Cadmium	0.50	+/-3	U	6.00	P	10/07/2025	19:24	LB137447
	Chromium	2.12	+/-5	U	10.0	P	10/07/2025	19:24	LB137447
	Lead	2.30	+/-6	U	12.0	P	10/07/2025	19:24	LB137447
	Selenium	9.64	+/-10	U	20.0	P	10/07/2025	19:24	LB137447
	Silver	1.62	+/-5	U	10.0	P	10/07/2025	19:24	LB137447