

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169971TB		WATER		Batch Number:	PB169997		Prep Date:	10/06/2025	
	Mercury	0.76	<2	U	2.00	CV	10/07/2025	11:41	LB137441
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169997BL		WATER		Batch Number:	PB169997		Prep Date:	10/06/2025	
	Mercury	0.076	<0.2	U	0.20	CV	10/07/2025	10:26	LB137441

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q3277

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169971TB		WATER		Batch Number:	PB169996		Prep Date:	10/06/2025	
	Arsenic	25.6	<50	U	100	P	10/07/2025	16:41	LB137447
	Barium	72.8	<250	U	500	P	10/07/2025	16:41	LB137447
	Cadmium	2.50	<15	U	30.0	P	10/07/2025	16:41	LB137447
	Chromium	10.6	<25	U	50.0	P	10/07/2025	16:41	LB137447
	Lead	11.5	<30	U	60.0	P	10/07/2025	16:41	LB137447
	Selenium	48.2	<50	U	100	P	10/07/2025	16:41	LB137447
	Silver	8.10	<25	U	50.0	P	10/07/2025	16:41	LB137447
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169996BL		WATER		Batch Number:	PB169996		Prep Date:	10/06/2025	
	Arsenic	25.6	<50	U	100	P	10/07/2025	18:11	LB137447
	Barium	72.8	<250	U	500	P	10/07/2025	18:11	LB137447
	Cadmium	2.50	<15	U	30.0	P	10/07/2025	18:11	LB137447
	Chromium	10.6	<25	U	50.0	P	10/07/2025	18:11	LB137447
	Lead	11.5	<30	U	60.0	P	10/07/2025	18:11	LB137447
	Selenium	48.2	<50	U	100	P	10/07/2025	18:11	LB137447
	Silver	8.10	<25	U	50.0	P	10/07/2025	18:11	LB137447