

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: G Environmental

SDG No.: Q3277

Contract: GENV01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	3.53			-20	20	10/07/2025	12:14	LB137447
	Barium	-8.30	6.0	138	-94	106	10/07/2025	12:14	LB137447
	Cadmium	2.40	1.0	240	-5	7	10/07/2025	12:14	LB137447
	Chromium	47.2	52.0	91	42	62	10/07/2025	12:14	LB137447
	Lead	-1.11			-12	12	10/07/2025	12:14	LB137447
	Selenium	8.06			-20	20	10/07/2025	12:14	LB137447
	Silver	-0.34			-10	10	10/07/2025	12:14	LB137447
ICSAB01	Arsenic	107	104	103	88.4	120	10/07/2025	12:37	LB137447
	Barium	477	537	89	437	637	10/07/2025	12:37	LB137447
	Cadmium	990	972	102	826	1120	10/07/2025	12:37	LB137447
	Chromium	540	542	100	460	624	10/07/2025	12:37	LB137447
	Lead	50.4	49.0	103	37	61	10/07/2025	12:37	LB137447
	Selenium	60.1	46.0	131	26	66	10/07/2025	12:37	LB137447
	Silver	205	201	102	170	232	10/07/2025	12:37	LB137447