

Continuing Calibration Report for SequenceID : FE100325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056163.D

Aliphatic C9-C12	8104359.000	60.000	3.311	6.944	135072.650	140027.957	3.539
Aliphatic C12-C16	6592546.000	40.000	6.945	10.393	164813.650	153376.491	-7.457
Aliphatic C16-C21	10831206.000	60.000	10.394	13.770	180520.100	165879.737	-8.826
Aliphatic C21-C28	12738583.000	80.000	13.771	17.439	159232.288	148527.151	-7.208
Aliphatic C28-C40	17718959.000	120.000	17.440	22.443	147657.992	136251.923	-8.371
Aliphatic EPH	55985653.000	360.000	3.311	22.443	155515.703	146449.789	-6.190