

Continuing Calibration Report for SequenceID : FE100325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056177.D

Aliphatic C9-C12	7448117.000	60.000	3.311	6.944	124135.283	140027.957	11.350
Aliphatic C12-C16	6053684.000	40.000	6.945	10.393	151342.100	153376.491	1.326
Aliphatic C16-C21	9852406.000	60.000	10.394	13.770	164206.767	165879.737	1.009
Aliphatic C21-C28	11632816.000	80.000	13.771	17.439	145410.200	148527.151	2.099
Aliphatic C28-C40	16324054.000	120.000	17.440	22.443	136033.783	136251.923	0.160
Aliphatic EPH	51311077.000	360.000	3.311	22.443	142530.769	146449.789	2.676