

Continuing Calibration Report for SequenceID : FE100325AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056190.D

Aliphatic C9-C12	7433078.000	60.000	3.311	6.944	123884.633	140027.957	11.529
Aliphatic C12-C16	6092853.000	40.000	6.945	10.393	152321.325	153376.491	0.688
Aliphatic C16-C21	9976890.000	60.000	10.394	13.770	166281.500	165879.737	-0.242
Aliphatic C21-C28	11763260.000	80.000	13.771	17.439	147040.750	148527.151	1.001
Aliphatic C28-C40	16330305.000	120.000	17.440	22.443	136085.875	136251.923	0.122
Aliphatic EPH	51596386.000	360.000	3.311	22.443	143323.294	146449.789	2.135