

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB092225

Review By	yogesh	Review On	9/24/2025 9:01:10 AM
Supervise By	mohammad	Supervise On	9/25/2025 2:24:54 AM
SubDirectory	FB092225	HP Acquire Method	FID_B
		HP Processing Method	FB092225
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24717,PP24931,PP24932,PP24933,PP24934,PP24935		
CCC	PP24937,PP24938		
Internal Standard/PEM			
ICV/I.BLK	PP24718,PP24936		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB032214.D	22 Sep 2025 9:53		YP/AJ	Ok
2	10 GRO STD		FB032215.D	22 Sep 2025 10:21		YP/AJ	Ok
3	20 GRO STD		FB032216.D	22 Sep 2025 10:48		YP/AJ	Ok
4	50 GRO STD		FB032217.D	22 Sep 2025 11:16		YP/AJ	Ok
5	100 GRO STD		FB032218.D	22 Sep 2025 11:44		YP/AJ	Ok
6	FB0922GROICV		FB032219.D	22 Sep 2025 12:43		YP/AJ	Ok
7	20 PPB GRO STD		FB032220.D	22 Sep 2025 13:23		YP/AJ	Ok
8	BSF0922S1		FB032221.D	22 Sep 2025 14:08	Not Used	YP/AJ	Not Ok
9	VBF0922S2		FB032222.D	22 Sep 2025 14:35		YP/AJ	Ok
10	VBF0922S1		FB032223.D	22 Sep 2025 15:03		YP/AJ	Ok
11	Q3123-01		FB032224.D	22 Sep 2025 15:31		YP/AJ	Ok
12	Q3157-01		FB032225.D	22 Sep 2025 15:59	Vial A not Purged	YP/AJ	Not Ok
13	BSF0922S1		FB032226.D	22 Sep 2025 16:35		YP/AJ	Ok
14	BSF0922S2		FB032227.D	22 Sep 2025 17:02		YP/AJ	Ok
15	20 PPB GRO STD		FB032228.D	22 Sep 2025 17:58		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB100825

Review By	yogesh	Review On	10/8/2025 3:11:07 PM
Supervise By	mohammad	Supervise On	10/11/2025 1:56:51 AM
SubDirectory	FB100825	HP Acquire Method	FID_B
		HP Processing Method	FB092225

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24717,PP24931,PP24932,PP24933,PP24934,PP24935
CCC Internal Standard/PEM	PP24991,PP24992,PP24993
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24718,PP24936

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB032236.D	8 Oct 2025 9:25		YP/AJ	Ok,M
2	VBF1008S1		FB032237.D	8 Oct 2025 10:11		YP/AJ	Ok
3	VBF1008S2		FB032238.D	8 Oct 2025 10:39		YP/AJ	Ok
4	BSF1008S1		FB032239.D	8 Oct 2025 11:06		YP/AJ	Ok
5	BSF1008S2		FB032240.D	8 Oct 2025 11:34		YP/AJ	Ok,M
6	Q3301-01		FB032241.D	8 Oct 2025 12:19	Vial A,Need 50X	YP/AJ	Dilution
7	Q3305-01		FB032242.D	8 Oct 2025 12:47	Carryover	YP/AJ	Not Ok
8	Q3277-01		FB032243.D	8 Oct 2025 13:15	Carryover	YP/AJ	Not Ok
9	Q3301-01		FB032244.D	8 Oct 2025 13:59	Vial C	YP/AJ	Ok
10	I.BLK		FB032245.D	8 Oct 2025 14:26		YP/AJ	Ok
11	20 PPB GRO STD		FB032246.D	8 Oct 2025 14:54		YP/AJ	Ok
12	Q3305-01		FB032247.D	8 Oct 2025 15:39	Vial B	YP/AJ	Ok
13	Q3305-01		FB032248.D	8 Oct 2025 16:06	Not Required	YP/AJ	Not Ok
14	Q3277-01		FB032249.D	8 Oct 2025 16:34	Vial B	YP/AJ	Ok
15	Q3277-01		FB032250.D	8 Oct 2025 17:01	Not Required	YP/AJ	Not Ok
16	BSF1008S3		FB032251.D	8 Oct 2025 17:29		YP/AJ	Ok
17	20 PPB GRO STD		FB032252.D	8 Oct 2025 18:24		YP/AJ	Ok

M : Manual Integration