

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3277

Continuing Calib Date: 10/06/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 09:49

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.74	4.74	4.64	4.84	0.00
Aroclor-1016-2	(2)	4.76	4.76	4.66	4.86	0.00
Aroclor-1016-3	(3)	4.81	4.82	4.72	4.92	0.01
Aroclor-1016-4	(4)	4.93	4.94	4.84	5.04	0.01
Aroclor-1016-5	(5)	5.19	5.19	5.09	5.29	0.00
Aroclor-1260-1	(1)	6.22	6.23	6.13	6.33	0.01
Aroclor-1260-2	(2)	6.41	6.42	6.32	6.52	0.01
Aroclor-1260-3	(3)	6.78	6.78	6.68	6.88	0.00
Aroclor-1260-4	(4)	7.03	7.04	6.94	7.14	0.01
Aroclor-1260-5	(5)	7.28	7.29	7.19	7.39	0.01
Tetrachloro-m-xylene		3.66	3.66	3.56	3.76	0.00
Decachlorobiphenyl		8.66	8.67	8.57	8.77	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: GENV01

Lab Code: ACE

SDG NO.: Q3277

Continuing Calib Date: 10/06/2025

Initial Calibration Date(s): 09/15/2025

09/15/2025

Continuing Calib Time: 09:49

Initial Calibration Time(s): 09:34

19:21

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.72	4.72	4.62	4.82	0.00
Aroclor-1016-2	(2)	4.73	4.74	4.64	4.84	0.01
Aroclor-1016-3	(3)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-4	(4)	4.95	4.96	4.86	5.06	0.01
Aroclor-1016-5	(5)	5.16	5.17	5.07	5.27	0.01
Aroclor-1260-1	(1)	6.19	6.20	6.10	6.30	0.01
Aroclor-1260-2	(2)	6.38	6.38	6.28	6.48	0.00
Aroclor-1260-3	(3)	6.53	6.54	6.44	6.64	0.01
Aroclor-1260-4	(4)	6.99	7.00	6.90	7.10	0.01
Aroclor-1260-5	(5)	7.24	7.25	7.15	7.35	0.01
Tetrachloro-m-xylene		3.65	3.65	3.55	3.75	0.01
Decachlorobiphenyl		8.60	8.61	8.51	8.71	0.01