

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q3277
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114249.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.737	4.643	4.843	589.600	500.000	17.9
Aroclor-1016-2	4.756	4.661	4.861	602.320	500.000	20.5
Aroclor-1016-3	4.813	4.718	4.918	593.270	500.000	18.7
Aroclor-1016-4	4.932	4.837	5.037	577.950	500.000	15.6
Aroclor-1016-5	5.188	5.093	5.293	611.650	500.000	22.3
Aroclor-1260-1	6.222	6.129	6.329	580.830	500.000	16.2
Aroclor-1260-2	6.412	6.319	6.519	569.840	500.000	14.0
Aroclor-1260-3	6.776	6.683	6.883	549.700	500.000	9.9
Aroclor-1260-4	7.034	6.941	7.141	644.310	500.000	28.9
Aroclor-1260-5	7.278	7.185	7.385	661.270	500.000	32.3
Decachlorobiphenyl	8.660	8.570	8.770	45.840	50.000	-8.3
Tetrachloro-m-xylene	3.657	3.560	3.760	51.480	50.000	3.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q3277
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114249.D **Time Analyzed:** 09:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.623	4.823	552.550	500.000	10.5
Aroclor-1016-2	4.733	4.641	4.841	560.670	500.000	12.1
Aroclor-1016-3	4.908	4.816	5.016	555.290	500.000	11.1
Aroclor-1016-4	4.950	4.858	5.058	561.850	500.000	12.4
Aroclor-1016-5	5.162	5.070	5.270	571.080	500.000	14.2
Aroclor-1260-1	6.187	6.096	6.296	532.050	500.000	6.4
Aroclor-1260-2	6.375	6.284	6.484	496.720	500.000	-0.7
Aroclor-1260-3	6.526	6.435	6.635	487.810	500.000	-2.4
Aroclor-1260-4	6.994	6.903	7.103	484.920	500.000	-3.0
Aroclor-1260-5	7.236	7.146	7.346	480.440	500.000	-3.9
Decachlorobiphenyl	8.603	8.514	8.714	45.990	50.000	-8.0
Tetrachloro-m-xylene	3.645	3.551	3.751	47.970	50.000	-4.1