

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q3277
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114264.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.643	4.843	582.670	500.000	16.5
Aroclor-1016-2	4.756	4.661	4.861	595.070	500.000	19.0
Aroclor-1016-3	4.813	4.718	4.918	569.650	500.000	13.9
Aroclor-1016-4	4.932	4.837	5.037	559.330	500.000	11.9
Aroclor-1016-5	5.187	5.093	5.293	586.500	500.000	17.3
Aroclor-1260-1	6.222	6.129	6.329	550.580	500.000	10.1
Aroclor-1260-2	6.412	6.319	6.519	535.800	500.000	7.2
Aroclor-1260-3	6.777	6.683	6.883	472.020	500.000	-5.6
Aroclor-1260-4	7.035	6.941	7.141	505.610	500.000	1.1
Aroclor-1260-5	7.279	7.185	7.385	509.360	500.000	1.9
Decachlorobiphenyl	8.661	8.570	8.770	37.070	50.000	-25.9
Tetrachloro-m-xylene	3.657	3.560	3.760	50.160	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** GENV01
Lab Code: ACE **SDG NO.:** Q3277
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/06/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114264.D **Time Analyzed:** 15:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.716	4.623	4.823	529.080	500.000	5.8
Aroclor-1016-2	4.735	4.641	4.841	554.850	500.000	11.0
Aroclor-1016-3	4.909	4.816	5.016	558.280	500.000	11.7
Aroclor-1016-4	4.952	4.858	5.058	560.240	500.000	12.0
Aroclor-1016-5	5.163	5.070	5.270	581.980	500.000	16.4
Aroclor-1260-1	6.189	6.096	6.296	488.920	500.000	-2.2
Aroclor-1260-2	6.377	6.284	6.484	458.970	500.000	-8.2
Aroclor-1260-3	6.528	6.435	6.635	445.750	500.000	-10.9
Aroclor-1260-4	6.996	6.903	7.103	443.550	500.000	-11.3
Aroclor-1260-5	7.239	7.146	7.346	436.000	500.000	-12.8
Decachlorobiphenyl	8.605	8.514	8.714	40.720	50.000	-18.6
Tetrachloro-m-xylene	3.647	3.551	3.751	49.050	50.000	-1.9