

Surrogate Summary

SDG No.: Q3277

Client: G Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO113719.D	PIBLK-PO113719.D	Tetrachloro-m-xyl	1	20	17.0	85		60	140
		Decachlorobiphen	1	20	17.3	86		60	140
		Tetrachloro-m-xyl	2	20	17.0	85		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
I.BLK-PO114253.D	PIBLK-PO114253.D	Tetrachloro-m-xyl	1	20	18.2	91		60	140
		Decachlorobiphen	1	20	17.6	88		60	140
		Tetrachloro-m-xyl	2	20	17.7	89		60	140
		Decachlorobiphen	2	20	18.1	91		60	140
PB169987BL	PB169987BL	Tetrachloro-m-xyl	1	20	20.5	102		32	144
		Decachlorobiphen	1	20	20.3	102		32	175
		Tetrachloro-m-xyl	2	20	19.4	97		32	144
		Decachlorobiphen	2	20	19.6	98		32	175
PB169987BS	PB169987BS	Tetrachloro-m-xyl	1	20	19.7	99		32	144
		Decachlorobiphen	1	20	20.1	101		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	19.9	100		32	175
Q3277-01	WC1	Tetrachloro-m-xyl	1	20	24.4	122		32	144
		Decachlorobiphen	1	20	22.6	113		32	175
		Tetrachloro-m-xyl	2	20	22.9	115		32	144
		Decachlorobiphen	2	20	22.9	115		32	175
Q3288-01MS	TR-01-100325MS	Tetrachloro-m-xyl	1	20	12.9	65		32	144
		Decachlorobiphen	1	20	13.8	69		32	175
		Tetrachloro-m-xyl	2	20	13.6	68		32	144
		Decachlorobiphen	2	20	14.6	73		32	175
Q3288-01MSD	TR-01-100325MSD	Tetrachloro-m-xyl	1	20	13.3	67		32	144
		Decachlorobiphen	1	20	12.7	64		32	175
		Tetrachloro-m-xyl	2	20	14.1	70		32	144
		Decachlorobiphen	2	20	14.1	70		32	175
I.BLK-PO114268.D	PIBLK-PO114268.D	Tetrachloro-m-xyl	1	20	17.7	88		60	140
		Decachlorobiphen	1	20	14.6	73		60	140
		Tetrachloro-m-xyl	2	20	17.5	87		60	140
		Decachlorobiphen	2	20	16.1	81		60	140