

Surrogate Summary

SDG No.: Q3277

Client: G Environmental

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3277-01	WC1	1,2-Dichloroethane-d4	50	64.3	129		63	155
		Dibromofluoromethane	50	55.5	111		70	134
		Toluene-d8	50	54.0	108		74	123
		4-Bromofluorobenzene	50	53.6	107		17	146
VW1003SBL01	VW1003SBL01	1,2-Dichloroethane-d4	50	62.5	125		63	155
		Dibromofluoromethane	50	53.4	107		70	134
		Toluene-d8	50	53.6	107		74	123
		4-Bromofluorobenzene	50	50.9	102		17	146
VW1003SBS01	VW1003SBS01	1,2-Dichloroethane-d4	50	50.6	101		63	155
		Dibromofluoromethane	50	49.6	99		70	134
		Toluene-d8	50	51.3	103		74	123
		4-Bromofluorobenzene	50	50.0	100		17	146
VW1003SBSD0	VW1003SBSD01	1,2-Dichloroethane-d4	50	51.2	102		63	155
		Dibromofluoromethane	50	51.3	103		70	134
		Toluene-d8	50	53.0	106		74	123
		4-Bromofluorobenzene	50	52.4	105		17	146