

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3277
 Lab File ID: VW032339.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	275706	7.96	492762	8.85	456443	11.63
UPPER LIMIT	551412	8.459	985524	9.349	912886	12.129
LOWER LIMIT	137853	7.459	246381	8.349	228222	11.129
EPA SAMPLE NO.						
WC1	196016	7.97	428903	8.86	434848	11.64
VW1003SBL01	187187	7.97	423770	8.85	435697	11.64
VW1003SBS01	278007	7.96	517265	8.85	472342	11.63
VW1003SBSD01	280986	7.96	514655	8.86	480184	11.64

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.