

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3277
 Lab File ID: VW032339.D Date Analyzed: 10/03/2025
 Instrument ID: MSVOA_W Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	225457	13.556				
UPPER LIMIT	450914	14.056				
LOWER LIMIT	112729	13.056				
EPA SAMPLE NO.						
WC1	211297	13.56				
VW1003SBL01	203854	13.56				
VW1003SBS01	234620	13.56				
VW1003SBSD01	244340	13.56				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.