

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q3278

Project: DPW

Run No.: LB137410

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137410BSW							
Bromide	mg/L	10	9.90		99	1	90-110	10/03/2025
Chloride	mg/L	3	2.90		97	1	90-110	10/03/2025
Fluoride	mg/L	2	2.00		100	1	90-110	10/03/2025
Nitrite	mg/L	3	3.00		100	1	90-110	10/03/2025
Nitrate	mg/L	2.5	2.40		96	1	90-110	10/03/2025
Sulfate	mg/L	15	14.8		99	1	90-110	10/03/2025
Orthophosphate as P	mg/L	5	4.60		92	1	90-110	10/03/2025