



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental

SDG No.: Q3278

Contract: GENV01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Iron	111	100	111	65 - 135	P	10/08/2025	12:41	LB137466
	Manganese	21.1	20.0	105	65 - 135	P	10/08/2025	12:41	LB137466
	Sodium	1840	2000	92	65 - 135	P	10/08/2025	12:41	LB137466
CRI01	Iron	112	100	112	65 - 135	P	10/09/2025	12:04	LB137484
	Manganese	21.1	20.0	106	65 - 135	P	10/09/2025	12:04	LB137484
	Sodium	1610	2000	80	65 - 135	P	10/09/2025	12:04	LB137484