

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3278</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	3080		1560		1500	102		P
Manganese	ug/L	75 - 125	126		30.1		100	96		P
Sodium	ug/L	75 - 125	381000		371000		1500	705		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>G Environmental</u>	level: <u>low</u>	sdg no.: <u>Q3278</u>
contract: <u>GENV01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3268-01</u>	client id: <u>WATER-TREATMENT-DISCHARGEMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3268-01MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	3070		1560		1500	100		P
Manganese	ug/L	75 - 125	126		30.1		100	96		P
Sodium	ug/L	75 - 125	376000		371000		1500	347		P