

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q3278

Contract: GENV01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169978BS							
Iron	ug/L	1500	1520		101	80 - 120	P
Manganese	ug/L	100	104		104	80 - 120	P
Sodium	ug/L	1500	1570		105	80 - 120	P