

Surrogate Summary

SDG No.: **Q3278**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3278-01	MW10	1,2-Dichloroethane-d4	50	47.2	94		70 (74)	130 (125)
		Dibromofluoromethane	50	47.2	94		70 (75)	130 (124)
		Toluene-d8	50	48.3	97		70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.6	111		70 (77)	130 (121)
VX1007WBL01	VX1007WBL01	1,2-Dichloroethane-d4	50	57.3	115		70 (74)	130 (125)
		Dibromofluoromethane	50	51.5	103		70 (75)	130 (124)
		Toluene-d8	50	49.3	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.8	114		70 (77)	130 (121)
VX1007WBS01	VX1007WBS01	1,2-Dichloroethane-d4	50	54.3	109		70 (74)	130 (125)
		Dibromofluoromethane	50	55.8	112		70 (75)	130 (124)
		Toluene-d8	50	53.5	107		70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.7	109		70 (77)	130 (121)
VX1007WBSD0	VX1007WBSD01	1,2-Dichloroethane-d4	50	53.0	106		70 (74)	130 (125)
		Dibromofluoromethane	50	54.9	110		70 (75)	130 (124)
		Toluene-d8	50	53.2	106		70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.5	107		70 (77)	130 (121)

