

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3278
 Lab File ID: VX048036.D Date Analyzed: 10/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144874	5.53	249421	6.74	221582	10.04
UPPER LIMIT	289748	6.025	498842	7.239	443164	10.537
LOWER LIMIT	72437	5.025	124711	6.239	110791	9.537
EPA SAMPLE NO.						
MW10	190887	5.54	376118	6.75	377293	10.04
VX1007WBL01	133238	5.54	276412	6.75	291549	10.04
VX1007WBS01	127551	5.53	226030	6.75	211007	10.04
VX1007WBSD01	129287	5.53	226163	6.75	207970	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.