

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GENV01
 Lab Code: ACE SDG NO.: Q3278
 Lab File ID: VX048036.D Date Analyzed: 10/07/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:44
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	111843	12.00€				
UPPER LIMIT	223686	12.50€				
LOWER LIMIT	55921.5	11.50€				
EPA SAMPLE NO.						
MW10	186931	12.01				
VX1007WBL01	144709	12.01				
VX1007WBS01	108312	12.01				
VX1007WBSD01	105986	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.