

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3279

Client ID : SSTD020247

Date Analyzed: 10/17/2025

Lab File ID: BN038092.D

Time Analyzed: 11:13

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	264714	7.781	1082740	10.59	667580	14.45
UPPER LIMIT	529428	8.281	2165480	11.087	1335160	14.945
LOWER LIMIT	132357	7.281	541370	10.087	333790	13.945
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	253439	7.78	1004480	10.59	621564	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.