

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3279

Client ID: SSTD020247

Date Analyzed: 10/17/2025

Lab File ID: BN038092.D

Time Analyzed: 11:13

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1176430	17.198	1146300	21.451	1132270	24.504
UPPER LIMIT	2352860	17.698	2292600	21.951	2264540	25.004
LOWER LIMIT	588215	16.698	573150	20.951	566135	24.004
EPA SAMPLE NO.						
SVOC-GPC2-BLANK	1121560	17.19	1026010	21.45	1016530	24.50

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.