

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038089.D
Instrument ID: BNA_N

Contract: CHEM02
SDG NO.: Q3279
DFTPP Injection Date: 10/17/2025
DFTPP Injection Time: 09:16

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	36.1
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	13.7
442	Greater than 50% of mass 198	92.2
443	15.0 - 24.0% of mass 442	18.2 (19.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SVOC-GPC2-BLANK	Q3279-08	BN038098.D	10/17/2025	16:37
SSTD020385	SSTDCCC020EC	BN038099.D	10/17/2025	17:16