

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>CHEM02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3279</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:23</u> <u>12:01</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D		
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.447	0.575	0.582	0.501	0.510	0.492	0.518	10
Chloromethane	0.610	0.749	0.739	0.661	0.677	0.647	0.680	7.9
Vinyl Chloride	0.572	0.713	0.716	0.661	0.677	0.658	0.666	7.9
Ethyl Acetate	0.429	0.561	0.554	0.532	0.533	0.539	0.525	9.2
Bromomethane		0.493	0.451	0.424	0.423	0.321	0.422	15
Chloroethane	0.433	0.450	0.469	0.443	0.450	0.426	0.445	3.4
Trichlorofluoromethane	0.865	1.033	1.038	0.995	1.018	0.986	0.989	6.5
1,1,2-Trichlorotrifluoroethane	0.519	0.591	0.603	0.583	0.590	0.583	0.578	5.2
Tert butyl alcohol		0.085	0.086	0.089	0.090	0.094	0.089	4.2
1,1-Dichloroethene	0.526	0.612	0.618	0.604	0.609	0.596	0.594	5.7
Acrolein		0.071	0.069	0.082	0.089	0.097	0.082	14.6
Acrylonitrile	0.264	0.329	0.343	0.347	0.343	0.344	0.328	9.8
Acetone	0.343	0.416	0.400	0.317	0.302	0.299	0.346	14.7
Carbon Disulfide	1.555	1.735	1.758	1.566	1.589	1.555	1.626	5.8
Methyl tert-butyl Ether	1.884	2.238	2.221	2.247	2.212	2.209	2.169	6.5
Methyl Acetate	0.612	0.673	0.714	0.876	0.865	0.881	0.770	15.4
Methylene Chloride	0.718	0.772	0.754	0.730	0.716	0.705	0.732	3.5
trans-1,2-Dichloroethene	0.543	0.703	0.665	0.649	0.643	0.635	0.640	8.3
Vinyl Acetate	1.541	1.937	2.020	1.992	1.958	1.961	1.901	9.4
1,1-Dichloroethane	1.090	1.325	1.299	1.297	1.287	1.278	1.263	6.8
Cyclohexane		1.161	1.092	1.030	0.984	0.993	1.052	7.1
2-Butanone	0.370	0.470	0.473	0.439	0.414	0.424	0.432	9
Carbon Tetrachloride	0.482	0.574	0.555	0.539	0.526	0.518	0.532	6
2,2-Dichloropropane	0.945	1.096	1.074	1.068	1.055	1.060	1.050	5.1
cis-1,2-Dichloroethene	0.698	0.806	0.803	0.808	0.787	0.797	0.783	5.4
Bromochloromethane	0.457	0.607	0.450	0.577	0.587	0.630	0.551	14.2
Chloroform	1.034	1.364	1.366	1.323	1.282	1.288	1.276	9.7
1,1,1-Trichloroethane	0.901	1.123	1.141	1.127	1.100	1.103	1.082	8.3
Methylcyclohexane	0.490	0.589	0.591	0.565	0.548	0.554	0.556	6.6
1,1-Dichloropropene	0.484	0.520	0.508	0.487	0.471	0.469	0.490	4.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>CHEM02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3279</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:23</u> <u>12:01</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D			
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
Benzene	1.330	1.577	1.566	1.523	1.473	1.459	1.488	6.1	
1,2-Dichloroethane	0.513	0.595	0.600	0.582	0.554	0.553	0.566	5.8	
Trichloroethene	0.296	0.382	0.382	0.370	0.364	0.363	0.360	8.9	
1,2-Dichloropropane	0.326	0.403	0.394	0.397	0.382	0.384	0.381	7.4	
Dibromomethane	0.248	0.285	0.289	0.288	0.276	0.277	0.277	5.5	
Bromodichloromethane	0.461	0.599	0.601	0.607	0.579	0.586	0.572	9.7	
4-Methyl-2-Pentanone	0.390	0.493	0.503	0.511	0.479	0.487	0.477	9.3	
Toluene	0.841	0.967	0.960	0.940	0.897	0.897	0.917	5.2	
t-1,3-Dichloropropene	0.487	0.588	0.602	0.619	0.595	0.611	0.584	8.3	
cis-1,3-Dichloropropene	0.513	0.652	0.644	0.653	0.637	0.644	0.624	8.7	
1,1,2-Trichloroethane	0.323	0.365	0.371	0.367	0.344	0.351	0.354	5.1	
1,3-Dichloropropane	0.507	0.648	0.636	0.638	0.607	0.609	0.607	8.5	
2-Chloroethyl Vinyl ether	0.094	0.212	0.210	0.272	0.280	0.292	0.227	32.7	
2-Hexanone	0.274	0.374	0.372	0.360	0.333	0.343	0.343	10.8	
Dibromochloromethane	0.321	0.414	0.431	0.435	0.421	0.422	0.407	10.5	
1,2-Dibromoethane	0.287	0.379	0.384	0.379	0.364	0.367	0.360	10.1	
Tetrachloroethene	0.299	0.354	0.342	0.326	0.320	0.314	0.326	6	
Chlorobenzene	0.998	1.182	1.177	1.178	1.139	1.142	1.136	6.2	
1,1,1,2-Tetrachloroethane	0.327	0.404	0.402	0.414	0.404	0.402	0.392	8.3	
Hexachloroethane	0.434	0.565	0.574	0.593	0.606	0.614	0.564	11.8	
Ethyl Benzene	1.625	2.077	2.049	2.039	1.996	1.973	1.960	8.6	
m/p-Xylenes	0.604	0.774	0.765	0.765	0.739	0.730	0.729	8.8	
o-Xylene	0.576	0.734	0.734	0.749	0.721	0.719	0.706	9.1	
Styrene	1.019	1.288	1.290	1.299	1.257	1.265	1.236	8.7	
Bromoform	0.221	0.293	0.301	0.313	0.312	0.317	0.293	12.4	
Isopropylbenzene	3.124	3.907	3.986	4.047	3.907	3.839	3.801	8.9	
1,1,2,2-Tetrachloroethane	0.965	1.182	1.203	1.225	1.152	1.173	1.150	8.2	
1,2,3-Trichloropropane	0.847	1.023	1.073	1.128	0.939	1.064	1.012	10.1	
Bromobenzene	0.760	0.964	0.992	0.988	0.960	0.954	0.936	9.4	
n-propylbenzene	3.680	4.667	4.739	4.792	4.581	4.504	4.494	9.2	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>Alliance</u>	Contract: <u>CHEM02</u>
Lab Code: <u>ACE</u>	SDG No.: <u>Q3279</u>
Instrument ID: <u>MSVOA_X</u>	Calibration Date(s): <u>09/16/2025</u> <u>09/16/2025</u>
Heated Purge: (Y/N) <u>N</u>	Calibration Time(s): <u>09:23</u> <u>12:01</u>
GC Column: <u>DB-624UI</u> ID: <u>0.18</u> (mm)	

LAB FILE ID:		RRF001 = VX047585.D		RRF005 = VX047586.D		RRF020 = VX047587.D		
		RRF050 = VX047588.D		RRF100 = VX047589.D		RRF150 = VX047590.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
2-Chlorotoluene	2.276	2.859	2.935	2.915	2.798	2.730	2.752	8.9
1,3,5-Trimethylbenzene	2.458	3.235	3.311	3.343	3.221	3.171	3.123	10.6
4-Chlorotoluene	2.754	3.347	3.410	3.433	3.285	3.275	3.251	7.7
tert-Butylbenzene	2.800	3.223	3.346	3.359	3.240	3.215	3.197	6.4
1,2,4-Trimethylbenzene	2.522	3.302	3.381	3.344	3.234	3.216	3.166	10.2
sec-Butylbenzene	3.110	4.001	3.945	4.028	3.867	3.834	3.797	9.1
p-Isopropyltoluene	2.660	3.372	3.371	3.419	3.263	3.205	3.215	8.8
1,3-Dichlorobenzene	1.462	1.800	1.829	1.817	1.768	1.758	1.739	8
1,4-Dichlorobenzene	1.570	1.890	1.868	1.832	1.767	1.783	1.785	6.5
n-Butylbenzene	2.429	3.070	3.114	3.139	3.098	2.999	2.975	9.1
1,2-Dichlorobenzene	1.359	1.691	1.757	1.746	1.696	1.692	1.657	9
1,2-Dibromo-3-Chloropropane	0.173	0.218	0.242	0.254	0.250	0.260	0.233	14.1
1,2,4-Trichlorobenzene	0.831	1.099	1.107	1.156	1.140	1.126	1.077	11.3
Hexachlorobutadiene	0.341	0.377	0.361	0.384	0.377	0.354	0.366	4.5
Naphthalene	2.503	3.153	3.410	3.539	3.496	3.571	3.279	12.5
1,2,3-Trichlorobenzene	0.753	0.986	1.031	1.088	1.078	1.073	1.002	12.7
1,2-Dichloroethane-d4		0.884	0.647	0.770	0.815	0.863	0.796	11.8
Dibromofluoromethane		0.356	0.266	0.331	0.355	0.368	0.335	12.3
Toluene-d8		1.237	0.943	1.147	1.211	1.263	1.160	11.1
4-Bromofluorobenzene		0.478	0.360	0.427	0.452	0.484	0.440	11.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.