

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG No.: Q3279

Instrument ID: MSVOA_X

Calibration Date/Time: 10/06/2025 08:54

Lab File ID: VX048009.D

Init. Calib. Date(s): 09/16/2025 09/16/2025

Heated Purge: (Y/N) N

Init. Calib. Time(s): 09:23 12:01

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.605		16.8	20
Chloromethane	0.680	0.636	0.1	-6.47	20
Vinyl Chloride	0.666	0.691		3.75	20
Ethyl Acetate	0.525	0.475		-9.52	20
Bromomethane	0.422	0.421		-0.24	20
Chloroethane	0.445	0.444		-0.22	20
Trichlorofluoromethane	0.989	1.092		10.41	20
1,1,2-Trichlorotrifluoroethane	0.578	0.657		13.67	20
Tert butyl alcohol	0.089	0.070		-21.35	20
1,1-Dichloroethene	0.594	0.601		1.18	20
Acrolein	0.082	0.087		6.1	20
Acrylonitrile	0.328	0.314		-4.27	20
Acetone	0.346	0.342		-1.16	20
Carbon Disulfide	1.626	1.517		-6.7	20
Methyl tert-butyl Ether	2.169	2.043		-5.81	20
Methyl Acetate	0.770	0.794		3.12	20
Methylene Chloride	0.732	0.697		-4.78	20
trans-1,2-Dichloroethene	0.640	0.623		-2.66	20
Vinyl Acetate	1.901	1.777		-6.52	20
1,1-Dichloroethane	1.263	1.250	0.1	-1.03	20
Cyclohexane	1.052	0.962		-8.56	20
2-Butanone	0.432	0.396		-8.33	20
Carbon Tetrachloride	0.532	0.573		7.71	20
2,2-Dichloropropane	1.050	1.050		0	20
cis-1,2-Dichloroethene	0.783	0.750		-4.22	20
Bromochloromethane	0.551	0.549		-0.36	20
Chloroform	1.276	1.253		-1.8	20
1,1,1-Trichloroethane	1.082	1.094		1.11	20
Methylcyclohexane	0.556	0.563		1.26	20
1,1-Dichloropropene	0.490	0.489		-0.2	20
Benzene	1.488	1.493		0.34	20
1,2-Dichloroethane	0.566	0.554		-2.12	20
Trichloroethene	0.360	0.368		2.22	20
1,2-Dichloropropane	0.381	0.395		3.67	20
Dibromomethane	0.277	0.275		-0.72	20
Bromodichloromethane	0.572	0.607		6.12	20
4-Methyl-2-Pentanone	0.477	0.457		-4.19	20
Toluene	0.917	0.910		-0.76	20

All other compounds must meet a minimum RRF of 0.010.

RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3279</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/06/2025 08:54</u>
Lab File ID:	<u>VX048009.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
t-1,3-Dichloropropene	0.584	0.576		-1.37	20
cis-1,3-Dichloropropene	0.624	0.627		0.48	20
1,1,2-Trichloroethane	0.354	0.355		0.28	20
1,3-Dichloropropane	0.607	0.606		-0.17	20
2-Chloroethyl Vinyl ether	0.227	0.278		22.47	20
2-Hexanone	0.343	0.319		-7	20
Dibromochloromethane	0.407	0.436		7.13	20
1,2-Dibromoethane	0.360	0.364		1.11	20
Tetrachloroethene	0.326	0.330		1.23	20
Chlorobenzene	1.136	1.153	0.3	1.5	20
1,1,1,2-Tetrachloroethane	0.392	0.421		7.4	20
Hexachloroethane	0.564	0.607		7.62	20
Ethyl Benzene	1.960	1.991		1.58	20
m/p-Xylenes	0.729	0.745		2.19	20
o-Xylene	0.706	0.712		0.85	20
Styrene	1.236	1.248		0.97	20
Bromoform	0.293	0.313	0.1	6.83	20
Isopropylbenzene	3.801	3.830		0.76	20
1,1,2,2-Tetrachloroethane	1.150	1.131	0.3	-1.65	20
1,2,3-Trichloropropane	1.012	0.916		-9.49	20
Bromobenzene	0.936	0.917		-2.03	20
n-propylbenzene	4.494	4.566		1.6	20
2-Chlorotoluene	2.752	2.717		-1.27	20
1,3,5-Trimethylbenzene	3.123	3.153		0.96	20
4-Chlorotoluene	3.251	3.209		-1.29	20
tert-Butylbenzene	3.197	3.174		-0.72	20
1,2,4-Trimethylbenzene	3.166	3.126		-1.26	20
sec-Butylbenzene	3.797	3.872		1.98	20
p-Isopropyltoluene	3.215	3.235		0.62	20
1,3-Dichlorobenzene	1.739	1.695		-2.53	20
1,4-Dichlorobenzene	1.785	1.698		-4.87	20
n-Butylbenzene	2.975	3.027		1.75	20
1,2-Dichlorobenzene	1.657	1.632		-1.51	20
1,2-Dibromo-3-Chloropropane	0.233	0.217		-6.87	20
1,2,4-Trichlorobenzene	1.077	1.011		-6.13	20
Hexachlorobutadiene	0.366	0.352		-3.83	20
Naphthalene	3.279	2.940		-10.34	20
1,2,3-Trichlorobenzene	1.002	0.945		-5.69	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CHEM02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3279</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/06/2025 08:54</u>
Lab File ID:	<u>VX048009.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
1,2-Dichloroethane-d4	0.796	0.715		-10.18	20
Dibromofluoromethane	0.335	0.352		5.07	20
Toluene-d8	1.160	1.165		0.43	20
4-Bromofluorobenzene	0.440	0.426		-3.18	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.