



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100625\
Data File : VX048008.D
Acq On : 06 Oct 2025 08:12
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1628

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	14030	PASS
75	95	30	60	52.4	35512	PASS
95	95	100	100	100.0	67717	PASS
96	95	5	9	6.4	4316	PASS
173	174	0.00	2	1.4	678	PASS
174	95	50	100	70.3	47587	PASS
175	174	5	9	7.7	3686	PASS
176	174	95	101	97.4	46371	PASS
177	176	5	9	6.4	2959	PASS