



Initial and Continuing Calibration Verification

Tully Environmental, Inc Q3281 **Client:** SDG No.:

Transfer Station-SPDES LB137476 **Project:** RunNo.:

| Analyte | | Units | Result | True Value | % Recovery | Acceptance Window (%R) | Analysis Date |
|--------------|------|-------|--------|------------|---------------|---------------------------|------------------|
| Sample ID: | ICV1 | | | | | | |
| Ammonia as N | | mg/L | 0.99 | 1 | 99 | 90-110 | 10/09/2025 |
| Sample ID: | CCV1 | | | | | | |
| Ammonia as N | | mg/L | 0.98 | 1 | 98 | 90-110 | 10/09/2025 |
| Sample ID: | CCV2 | | | | | | |
| Ammonia as N | | mg/L | 1 | 1 | 100 | 90-110 | 10/09/2025 |
| Sample ID: | CCV3 | | | | | | |
| Ammonia as N | | mg/L | 0.99 | 1 | 99 | 90-110 | 10/09/2025 |