

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Janvi	Review On	10/8/2025 2:41:18 PM
Supervise By	jaswal	Supervise On	10/8/2025 11:43:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87074,MP87083,MP87082,MP87080,MP87079,MP87078,MP87077,MP87076,MP87075		
ICV Standard	MP87085,MP87082		
CCV Standard	MP87086		
ICSA Standard	MP87087,MP87088		
CRI Standard	MP87082		
LCS Standard			
Chk Standard	MP87091,MP87092		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/07/25 10:29		Jaswal	OK
2	S1	S1	CAL2	10/07/25 10:34		Jaswal	OK
3	S2	S2	CAL3	10/07/25 10:38		Jaswal	OK
4	S3	S3	CAL4	10/07/25 10:42		Jaswal	OK
5	S4	S4	CAL5	10/07/25 10:46		Jaswal	OK
6	S5	S5	CAL6	10/07/25 10:50		Jaswal	OK
7	ICV01	ICV01	ICV	10/07/25 10:54		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/07/25 12:01		Jaswal	OK
9	ICB01	ICB01	ICB	10/07/25 12:06		Jaswal	OK
10	CRI01	CRI01	CRDL	10/07/25 12:10		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/07/25 12:14		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/07/25 12:37		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/07/25 12:41		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/07/25 12:45		Jaswal	OK
15	CCV01	CCV01	CCV	10/07/25 12:50		Jaswal	OK
16	CCB01	CCB01	CCB	10/07/25 12:58		Jaswal	OK
17	Q3283-01	001-WILLETS-PT-BL	SAM	10/07/25 13:21		Jaswal	OK
18	Q3283-02	002-35TH-AVE(SEP)	SAM	10/07/25 13:25		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Janvi	Review On	10/8/2025 2:41:18 PM
Supervise By	jaswal	Supervise On	10/8/2025 11:43:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87074,MP87083,MP87082,MP87080,MP87079,MP87078,MP87077,MP87076,MP87075		
ICV Standard	MP87085,MP87082		
CCV Standard	MP87086		
ICSA Standard	MP87087,MP87088		
CRI Standard	MP87082		
LCS Standard			
Chk Standard	MP87091,MP87092		

19	Q3283-02DUP	002-35TH-AVE(SEP)DUP		10/07/25 13:29		Jaswal	OK
20	Q3283-02L	002-35TH-AVE(SEP)LS		10/07/25 13:33		Jaswal	OK
21	Q3283-02MS	002-35TH-AVE(SEP)MS		10/07/25 13:38		Jaswal	OK
22	Q3283-02MSD	002-35TH-AVE(SEP)MSD		10/07/25 13:42		Jaswal	OK
23	Q3283-02A	002-35TH-AVE(SEP)PS		10/07/25 13:46		Jaswal	OK
24	LR-1	LR-1	HIGH STD	10/07/25 13:50		Jaswal	OK
25	LR-3	LR-3	HIGH STD	10/07/25 13:59		Jaswal	OK
26	LR-2	LR-2	HIGH STD	10/07/25 14:10		Jaswal	OK
27	CCV02	CCV02	CCV	10/07/25 14:19		Jaswal	OK
28	CCB02	CCB02	CCB	10/07/25 14:24		Jaswal	OK
29	Q3270-01	SG-1-TEST-PIT	SAM	10/07/25 14:28		Jaswal	OK
30	Q3271-01	72-11929	SAM	10/07/25 14:32		Jaswal	OK
31	Q3271-01L	72-11929L	SD	10/07/25 14:40		Jaswal	OK
32	Q3271-01MS	72-11929MS	MS	10/07/25 14:44		Jaswal	OK
33	Q3271-01MSD	72-11929MSD	MSD	10/07/25 14:48		Jaswal	OK
34	Q3271-01A	72-11929A	PS	10/07/25 14:52		Jaswal	OK
35	Q3272-01	VNJ-235	SAM	10/07/25 14:56		Jaswal	OK
36	Q3272-03	R7	SAM	10/07/25 15:00	Ca is high, need dilution	Jaswal	Dilution
37	Q3277-01	WC1	SAM	10/07/25 15:04		Jaswal	OK
38	CCV03	CCV03	CCV	10/07/25 15:08		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Janvi	Review On	10/8/2025 2:41:18 PM
Supervise By	jaswal	Supervise On	10/8/2025 11:43:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87074,MP87083,MP87082,MP87080,MP87079,MP87078,MP87077,MP87076,MP87075		
ICV Standard	MP87085,MP87082		
CCV Standard	MP87086		
ICSA Standard	MP87087,MP87088		
CRI Standard	MP87082		
LCS Standard			
Chk Standard	MP87091,MP87092		

39	CCB03	CCB03	CCB	10/07/25 15:12		Jaswal	OK
40	Q3269-03	EG1-TP1	SAM	10/07/25 15:17		Jaswal	OK
41	Q3269-06	EG1-TP2	SAM	10/07/25 15:21		Jaswal	OK
42	Q3281-01	001-WILLETS-PT-BL	SAM	10/07/25 15:26		Jaswal	OK
43	Q3281-02	002-35TH-AVE(OCT)	SAM	10/07/25 15:30		Jaswal	OK
44	Q3272-03DL	R7DL	SAM	10/07/25 15:34	25X for Ca	Jaswal	Confirms
45	Q3271-01DUP	72-11929DUP	DUP	10/07/25 15:38		Jaswal	OK
46	PB169982BL	PB169982BL	MB	10/07/25 15:43		Jaswal	OK
47	PB169982BS	PB169982BS	LCS	10/07/25 15:48		Jaswal	OK
48	PB169979BL	PB169979BL	MB	10/07/25 15:52		Jaswal	OK
49	PB169979BS	PB169979BS	LCS	10/07/25 15:56		Jaswal	OK
50	CCV04	CCV04	CCV	10/07/25 16:00		Jaswal	OK
51	CCB04	CCB04	CCB	10/07/25 16:04		Jaswal	OK
52	PB169996BS	PB169996BS	LCS	10/07/25 16:08		Jaswal	OK
53	Q3266-02	ENV-4-6FT	SAM	10/07/25 16:13		Jaswal	OK
54	Q3266-02DUP	ENV-4-6FTDUP	DUP	10/07/25 16:17		Jaswal	OK
55	Q3266-02L	ENV-4-6FTL	SD	10/07/25 16:20		Jaswal	OK
56	Q3266-02MS	ENV-4-6FTMS	MS	10/07/25 16:25		Jaswal	OK
57	Q3266-02MSD	ENV-4-6FTMSD	MSD	10/07/25 16:29		Jaswal	OK
58	Q3266-02A	ENV-4-6FTA	PS	10/07/25 16:33		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Janvi	Review On	10/8/2025 2:41:18 PM
Supervise By	jaswal	Supervise On	10/8/2025 11:43:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87074,MP87083,MP87082,MP87080,MP87079,MP87078,MP87077,MP87076,MP87075		
ICV Standard	MP87085,MP87082		
CCV Standard	MP87086		
ICSA Standard	MP87087,MP87088		
CRI Standard	MP87082		
LCS Standard			
Chk Standard	MP87091,MP87092		

59	Q3258-06DL	COMPOSITEDL	SAM	10/07/25 16:37	Not use	Jaswal	Not Ok
60	PB169971TB	PB169971TB	MB	10/07/25 16:41		Jaswal	OK
61	Q3262-03	MOO-25-0280	SAM	10/07/25 16:45		Jaswal	OK
62	CCV05	CCV05	CCV	10/07/25 16:50		Jaswal	OK
63	CCB05	CCB05	CCB	10/07/25 17:01		Jaswal	OK
64	Q3262-06	MOO-25-0281-82	SAM	10/07/25 17:05		Jaswal	OK
65	Q3262-09	279887	SAM	10/07/25 17:10		Jaswal	OK
66	Q3277-02	WC1	SAM	10/07/25 17:14		Jaswal	OK
67	Q3289-02	VNJ-240	SAM	10/07/25 17:19		Jaswal	OK
68	Q3289-04	VNJ-261	SAM	10/07/25 17:23		Jaswal	OK
69	Q3289-06	72-11966	SAM	10/07/25 17:27		Jaswal	OK
70	Q3289-06DUP	72-11966DUP	DUP	10/07/25 17:32		Jaswal	OK
71	Q3289-06L	72-11966L	SD	10/07/25 17:36		Jaswal	OK
72	Q3289-06MS	72-11966MS	MS	10/07/25 17:40		Jaswal	OK
73	Q3289-06MSD	72-11966MSD	MSD	10/07/25 17:44		Jaswal	OK
74	CCV06	CCV06	CCV	10/07/25 17:49		Jaswal	OK
75	CCB06	CCB06	CCB	10/07/25 17:59		Jaswal	OK
76	Q3289-06A	72-11966A	PS	10/07/25 18:03		Jaswal	OK
77	Q3287-02	TSS-25-0001-0002	SAM	10/07/25 18:07		Jaswal	OK
78	PB169996BL	PB169996BL	MB	10/07/25 18:11		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137447

Review By	Janvi	Review On	10/8/2025 2:41:18 PM
Supervise By	jaswal	Supervise On	10/8/2025 11:43:16 PM
STD. NAME	STD REF.#		
ICAL Standard	MP87074,MP87083,MP87082,MP87080,MP87079,MP87078,MP87077,MP87076,MP87075		
ICV Standard	MP87085,MP87082		
CCV Standard	MP87086		
ICSA Standard	MP87087,MP87088		
CRI Standard	MP87082		
LCS Standard			
Chk Standard	MP87091,MP87092		

79	PB169983BL	PB169983BL	MB	10/07/25 18:15		Jaswal	OK
80	PB169983BS	PB169983BS	LCS	10/07/25 18:20		Jaswal	OK
81	Q3288-01	TR-01-100325	SAM	10/07/25 18:24		Jaswal	OK
82	Q3289-01	VNJ-240	SAM	10/07/25 18:28		Jaswal	OK
83	Q3289-03	VNJ-261	SAM	10/07/25 18:32		Jaswal	OK
84	Q3289-05	72-11966	SAM	10/07/25 18:36		Jaswal	OK
85	Q3290-01	CL-02-100325	SAM	10/07/25 18:40		Jaswal	OK
86	CCV07	CCV07	CCV	10/07/25 18:44		Jaswal	OK
87	CCB07	CCB07	CCB	10/07/25 18:48		Jaswal	OK
88	Q3290-01DUP	CL-02-100325DUP	DUP	10/07/25 18:52		Jaswal	OK
89	Q3290-01L	CL-02-100325L	SD	10/07/25 18:56		Jaswal	OK
90	Q3290-01MS	CL-02-100325MS	MS	10/07/25 19:00		Jaswal	OK
91	Q3290-01MSD	CL-02-100325MSD	MSD	10/07/25 19:04		Jaswal	OK
92	Q3290-01A	CL-02-100325A	PS	10/07/25 19:08		Jaswal	OK
93	Q3257-01DL	ENV1-6FTDL	SAM	10/07/25 19:12	5X for Ag	Jaswal	Confirms
94	Q3257-02DL	ENV2-6FTDL	SAM	10/07/25 19:16	5X for Ag	Jaswal	Confirms
95	CCV08	CCV08	CCV	10/07/25 19:20		Jaswal	OK
96	CCB08	CCB08	CCB	10/07/25 19:24		Jaswal	OK