

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3281

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	1020	1000	102	95 - 105	P	10/07/2025	10:54	LB137447
	Iron	4130	4000	103	95 - 105	P	10/07/2025	10:54	LB137447
	Lead	3890	4000	97	95 - 105	P	10/07/2025	10:54	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3281

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.9	20.0	110	80 - 120	P	10/07/2025	12:01	LB137447
	Iron	119	100	119	80 - 120	P	10/07/2025	12:01	LB137447
	Lead	11.1	12.0	93	80 - 120	P	10/07/2025	12:01	LB137447

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3281

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1270	1250	101	90 - 110	P	10/07/2025	12:50	LB137447
	Iron	5200	5000	104	90 - 110	P	10/07/2025	12:50	LB137447
	Lead	4980	5000	100	90 - 110	P	10/07/2025	12:50	LB137447
CCV02	Copper	1260	1250	101	90 - 110	P	10/07/2025	14:19	LB137447
	Iron	5340	5000	107	90 - 110	P	10/07/2025	14:19	LB137447
	Lead	4960	5000	99	90 - 110	P	10/07/2025	14:19	LB137447
CCV03	Copper	1240	1250	99	90 - 110	P	10/07/2025	15:08	LB137447
	Iron	5120	5000	102	90 - 110	P	10/07/2025	15:08	LB137447
	Lead	4840	5000	97	90 - 110	P	10/07/2025	15:08	LB137447
CCV04	Copper	1260	1250	101	90 - 110	P	10/07/2025	16:00	LB137447
	Iron	5080	5000	102	90 - 110	P	10/07/2025	16:00	LB137447
	Lead	4950	5000	99	90 - 110	P	10/07/2025	16:00	LB137447
CCV05	Copper	1250	1250	100	90 - 110	P	10/07/2025	16:50	LB137447
	Iron	5320	5000	106	90 - 110	P	10/07/2025	16:50	LB137447
	Lead	4890	5000	98	90 - 110	P	10/07/2025	16:50	LB137447
CCV06	Copper	1250	1250	100	90 - 110	P	10/07/2025	17:49	LB137447
	Iron	4840	5000	97	90 - 110	P	10/07/2025	17:49	LB137447
	Lead	4880	5000	98	90 - 110	P	10/07/2025	17:49	LB137447
CCV07	Copper	1260	1250	101	90 - 110	P	10/07/2025	18:44	LB137447
	Iron	4950	5000	99	90 - 110	P	10/07/2025	18:44	LB137447
	Lead	4920	5000	98	90 - 110	P	10/07/2025	18:44	LB137447
CCV08	Copper	1210	1250	96	90 - 110	P	10/07/2025	19:20	LB137447
	Iron	4810	5000	96	90 - 110	P	10/07/2025	19:20	LB137447
	Lead	4720	5000	94	90 - 110	P	10/07/2025	19:20	LB137447