

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q3281

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	13.8	2.0	690	-18	22	10/07/2025	12:14	LB137447
	Iron	94400	101000	94	85600	116500	10/07/2025	12:14	LB137447
	Lead	-1.11			-12	12	10/07/2025	12:14	LB137447
ICSAB01	Copper	489	511	96	434	588	10/07/2025	12:37	LB137447
	Iron	97100	99300	98	84400	114500	10/07/2025	12:37	LB137447
	Lead	50.4	49.0	103	37	61	10/07/2025	12:37	LB137447
ICSA	Copper	18.9	2.0	945	-18	22	10/07/2025	12:41	LB137447
	Iron	99800	101000	99	85600	116500	10/07/2025	12:41	LB137447
	Lead	-1.50			-12	12	10/07/2025	12:41	LB137447
ICSAB	Copper	521	511	102	434	588	10/07/2025	12:45	LB137447
	Iron	93400	99300	94	84400	114500	10/07/2025	12:45	LB137447
	Lead	17.4	49.0	36	37	61	10/07/2025	12:45	LB137447