

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TULL01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3281</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>10/08/2025</u> <u>10:08</u>
Lab File ID:	<u>VN088031.D</u>	Init. Calib. Date(s):	<u>10/08/2025</u> <u>10/08/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:34</u> <u>11:12</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.426	1.402		-1.68	
Toluene	1.702	1.700		-0.12	
Ethyl Benzene	1.850	1.839		-0.6	
m/p-Xylenes	0.713	0.714		0.14	
o-Xylene	0.710	0.710		0	
1,2-Dichloroethane-d4	2.336	2.319		-0.73	
Toluene-d8	1.337	1.343		0.45	
4-Bromofluorobenzene	0.495	0.497		0.4	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.