

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
Lab Code: ACE SDG NO.: Q3281
Lab File ID: VN088031.D Date Analyzed: 10/08/2025
Instrument ID: MSVOA_N Time Analyzed: 10:08
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	71387	7.79	401663	9.08	355387	11.85
UPPER LIMIT	142774	8.294	803326	9.582	710774	12.347
LOWER LIMIT	35693.5	7.294	200832	8.582	177694	11.347
EPA SAMPLE NO.						
001-WILLETTS-PT-BLVD (OCT)	64341	7.79	356991	9.08	318567	11.85
002-35TH-AVE (OCT)	64155	7.80	352319	9.08	309016	11.85
VN1008WBL01	61923	7.80	352790	9.08	309337	11.84
VN1008WBS01	68450	7.80	383696	9.08	341991	11.85
VN1008WBSD01	67745	7.79	378996	9.08	337371	11.85

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.