

Report of Analysis

Client:	Tully Environmental, Inc	Date Collected:	10/02/25
Project:	Transfer Station-SPDES	Date Received:	10/03/25
Client Sample ID:	001-WILLETS-PT-BLVD(SEP)	SDG No.:	Q3283
Lab Sample ID:	Q3283-01	Matrix:	Water
Level (low/med):	low	% Solid:	0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	1.97	J	1	1.89	10.0	ug/L	10/06/25 09:20	10/07/25 13:21	EPA 200.7	
7439-89-6	Iron	758		1	18.0	50.0	ug/L	10/06/25 09:20	10/07/25 13:21	EPA 200.7	
7439-92-1	Lead	2.92	J	1	1.21	6.00	ug/L	10/06/25 09:20	10/07/25 13:21	EPA 200.7	

Color Before:	Colorless	Clarity Before:	Clear	Texture:
Color After:	Colorless	Clarity After:	Clear	Artifacts:
Comments:	Metals Group 10			

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 D = Dilution
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 * = indicates the duplicate analysis is not within control limits.
 E = Indicates the reported value is estimated because of the presence of interference.
 OR = Over Range
 N =Spiked sample recovery not within control limits